



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340308/319
E mail: procurement@wbmsc.gov.in

PROCUREMENT OF MEDICAL EQUIPMENT FOR MEDICAL COLLEGE AND HOSPITALS
(GROUP 3)

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT-217/2018

Dated –12.10.2018

The following amendment has been made in the tender document,

Amendment – 2

Schedule V

HYSTEROSCOPE MACHINE :

PARAMETER / FEATURE	THRESHOLD LIMIT / RANGE
FORWARD-OBLIQUE TELESCOPE IN DEGREE	30
Magnification	1X,60X
DIAMETER in mm	4.0
LENGTH in cm	30

EXAMINATION SHEATH DIAMETER	5.0-5.1 mm TO BE USED WITH 4mm TELESCOPE WITH 1 LUER-LOCK ADAPTOR
INNER SIZE DIAMETER	5.2-5.3 mm & OUTER DIA SIZE 6.2-6.3 mm TO BE USED WITH 5 FR CHANNEL WITH 1 STOPCOCK AND 1 LUER-LOCK ADAPTOR
SCISSOR	5FR, LENGTH 34 cm
SCISSORS POINTED JAWS	5 FR, LENGTH 34 cm
NEEDLE ELECTRODE	5 FR, LENGTH 27 cm
BIPOLAR VAPORISATION ELECTRODE	5 FR, LENGTH 36 cm
HIGH FREQUENCY CORD	4mm PLUG HF-UNIT, 300cm
MUST HAVE SPECIFICATION	
PARAMETER / FEATURE	YES / NO
OPERATING AND CONTRACT-HYSTEROSCOPE FORWARD OBLIQUE TELESCOPE	
AUTOCLAVABLE	
ENLARGED VIEW	
FIBER OPTIC LIGHT TRANSMISSION INCORPORATED	
CONTINUOUS FLOW EXAMINATION SHEATH	
OBTURATOR FOR USE WITH THE SHEATHS	
SCISSORS SHOULD HAVE BLUNT TIPS AND SINGLE ACTION JAWS	
SCISSORS POINTED JAWS SINGLE ACTIONS JAWS SEMI RIGID	
NEEDLE ELECTRODE WITH UNIPOLAR /BIOPOLAR HIGH FREQUENCY CORD	
HYSTEROMAT	
WORKING ELEMENT(RESECTOSCOPE) WITH ELECTRODES	
LOOP,BALL,NEEDLE FOR TRANSCERVICAL RESECTION OF ENDOMETRIUM(TCRE)	
BIOPSY-AND GRASPING FORCEPS WITH DOUBLE ACTION JAWS	
QUALITY STANDARD AND SAFETY CERTIFICATION	
EUROPIAN CE / US FDA	