

# Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited  
Swasthya Sathi  
GN-29, Salt Lake, Sector-V  
Kolkata-700091

Phone No (033) 40340308/319  
E mail: [procurement@wbmsc.gov.in](mailto:procurement@wbmsc.gov.in)

SUPPLY AND COMMISSIONING OF BLOOD-BANK EQUIPMENTS IN THE HOSPITALS AND MEDICAL COLLEGES OF THE GOVERNMENT OF WEST BENGAL.

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT- 56/2016

Dated – 29.09.2016

## AMENDMENT - I

### TECHNICAL SPECIFICATIONS

#### SCHEDULE - II

#### CENTRIFUGE (TABLE TOP)

- a) Capacity: At least 16 holes x15ml round bottom test tube (plastic & glass).
- b) Built in time and speed regulator (digital controlled) with suitable speed & time indicator (digital display) and lid lock system.
- c) RPM: 500 to 4000 rpm with full load with swing out rotor head.
- d) Power Supply: 220/240 volts, single phase, 50 Hz
- e) A line voltage corrector of appropriate rating will form part of standard configuration. The servo voltage corrector should be capable to correct input voltage range from 160 – 280 V AC to 220/240 VAC, 50Hz.
- f) Technical Literature: The firm shall submit printed illustrated technical literature/ leaflet indicating the model number. If quoted model is a modified version of their any standard product that also be indicated in the offer
- g) Product certification: CE / US FDA / BIS certified.
- h) Quality Certification: ISO certified.