

# Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited  
Swasthya Sathi  
GN-29, Salt Lake, Sector-V  
Kolkata-700091

Phone No (033) 40340307/40340308

E mail: [procurement@wbmsc.gov.in](mailto:procurement@wbmsc.gov.in)

SUPPLY OF MEDICAL EQUIPMENTS FOR MULTI / SUPER SPECIALITY HOSPITALS OF THE GOVERNMENT OF WEST BENGAL.

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT-15/2016

Dated – 14.04.2016

The following amendments have been made in the tender document,

## Amendment - I

### REVISED TECHNICAL SPECIFICATION

#### SCHEDULE - X

#### Hot Air oven

1. The entire system should be Double Walled, inner wall & outer wall should be SS-304.
2. Shelves: 2 to 3 stainless steel, perforated shelves, could be adjusted at any levels.
3. Heating Elements: placed at the bottom and sides.
4. Temperature control: Thermostatic control from 40 to 200° C with Regulator and digital display temperature-indicator.
5. Microprocessor PID Controller.
6. Accuracy Range:  $\pm 2^{\circ}\text{C}$ .
7. Should have the provision for door alarm & self diagnostic alarm.
8. Should be supplied with one Centigrade Thermometer fitted with the equipment for cross verification of the digitally measured temperature.
9. Thermometer should be calibrated along with calibration certificate.
10. Automatic timer should be at least 60 minutes.
11. Air circulating fan.
12. Volume : 50 - 70 Litre
13. Standard, Safety and Training
14. Product should be CE certified.
15. Documentation
  - i. User Manual & Service Manual should be in English