

Notice Inviting e-Tender
West Bengal Medical Services Corporation Limited
Swasthya Bhawan Complex
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 23577070/40340307/40340308

E mail: deo.wbmisc@gmail.com, md.wbmisc@gmail.com

SUPPLY OF BLOOD-BANK EQUIPMENTS IN THE HOSPITALS AND MEDICAL COLLEGES OF THE GOVERNMENT OF WEST BENGAL.

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT-60/2014

Dated- 31.10.2014

The following amendment have been made in the tender document,

Amendment - I

REVISED TECHNICAL SPECIFICATION

Schedule II

CENTRIFUGE (TABLE TOP)

- a) Capacity: **At least** 16 holes x15ml **round bottom test tube.**
- b) Built in time and speed regulator (digital controlled) with suitable speed indication (digital display) and lid lock system
- c) RPM: 500 to **4500 rpm with full load with swing out rotor head.**
- d) Power Supply: 220/240 volts, single phase, 50 Hz
- e) A line voltage corrector of suitable rating from reputed brand should be supplied with the equipment
- f) Technical Literature: The firm shall submit printed illustrated technical literature/ leaflet indicating the model number. If quoted model is a modified version of their any standard product that also be indicated in the offer
- g) Product certification: CE / US FDA / BIS certified.
- h) Quality Certification: ISO certified.