



## Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited  
Swasthya Sathi  
GN-29, Salt Lake, Sector-V  
Kolkata-700091

Phone No (033) 40340307/412  
E mail: [procurement@wbmsc.gov.in](mailto:procurement@wbmsc.gov.in)

SUPPLY AND INSTALLATION OF DIFFERENT TYPES OF MEDICAL EQUIPMENT IN THE  
HOSPITALS AND MEDICAL COLLEGES OF THE GOVERNMENT OF WEST BENGAL.  
(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT-464/2023

Dated – 09.08.2023

The following amendment have been made in the tender document,

### Amendment –I (Revision of Technical Specification)

The revised technical specifications for the item is given below,

#### Phase-contrast microscope with photographic attachment

Optics:	Should have Total Infinity Corrected Optical system / Harmonic Component System (HCS)
Tube:	Binocular tube with 45 deg viewing angle
Magnification:	40X - 400X
Eyepieces:	10X adjustable eyepiece with FOV 20mm
IPD:	Automatic constant sharpness from eye width of at least 55 - 75mm
Nosepiece:	Revolving quadruple nosepiece
Stage:	Should have 3 position s stage support with object guide for X -Y travel of the sample (adjustment range 128 x 83 mm), with holder for petridish, slides, flask etc.
Illumination:	Precentred at least 5W Long life LED illumination, with constant color temperature, should have automatic brightness adjustment to contrast method & with life time and auto-off function for illumination.

Condenser: Long free condenser of working distance of 40 mm and a numerical aperture of 0.45. Specially made for Bright field & Phase Contrast.

Objectives: Universal Objective for Bright field & Phase contrast

PLAN ACHROMAT 4x/0.10 PH0 (Free working distance: 18.0 mm)

PLAN ACHROMAT 10x/0.22 PH1 (Free working distance: 7.8 mm)

PLAN ACHROMAT 20x/0.30 PH1 (Free working distance: 3.7-2.4 mm)

PLAN ACHROMAT 40x/0.50 PH2 (Free working distance: 2.0 mm)

Digital camera: Digital camera with sw kit, colour digital camera for on-screen microscope image display, resolution (No. of pixels) 2048x1536, 10 to 12 Mpixels, CMOS sensor progressive scan, pixel size 1.5  $\mu$ m x 1.5  $\mu$ m, colour filter RGB (Bayer), colour depth 30 bits, exposure time 0.1 msec- 2 sec, A/D converter 10 bit, dynamic range >45dB, full size live image, 'max. 15 frames per sec. VGA, SVGA, XGA and XVGA image formats, gain 1x - 20x, shading correction, optical C-mount, recommended video adapter 0.5x, latest supported operating systems with branded PC.

Product Certification: European CE (4 digit notify body)/ European Conformity for quoted model / US FDA/ BIS / CDSCO