



## Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited  
Swasthya Sathi  
GN-29, Salt Lake, Sector-V  
Kolkata-700091

Phone No (033) 40340308/319  
E mail: procurement@wbmsc.gov.in

SUPPLY OF EQUIPMENTS FOR CHRONIC OBSTRUCTIVE PULMONARY DISEASE (COPD) IN THE HOSPITALS AND MEDICAL COLLEGES OF THE GOVERNMENT OF WEST BENGAL.

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT-16/2020

Dated –10.02.2020

The following amendments have been made in the tender document,

### Amendment –I (Revision of Technical Specification)

The revised technical specifications for the item is given below,

### Portable Spiro meter with non thermal printing system

1. Should measure FVC, SVC and MVV test.
2. Should be able to evaluate 51 vital parameters with interpretation. The following parameters should be measured by the PFT Equipment. It should measure FEV1 , FVC, FEV1/FVC ratio, PEF, SVC, MMEF, FEF25%-75%, reversibility , PIF, MVV, FET, ERV, IRV
3. Predicted value- depends upon national preference. It should provide Indian predicted equations.
4. Multi window lay out.
5. Configurable print out format.

6. Real time flow volume and volume time traces.
7. Pre and post test facility with graphical display of pre and post- bronchodilator tests separately.
8. Overlaying of previous trial curves for comparison.
9. Open & Closed flow/volume looptest technique possible.
10. Storage- 2500 patients' tests including flow/volume loops and volume time curves.
11. Should have networking support.
12. Should be provided with disposable mouthpieces – 1000 nos.
13. Ultrasonic sensor for PFT.
14. Should have high resolution colour graphic TFT.
15. Should have CE or US FDA approval.
16. Inkjet printer should be supply with the machine.
17. Banded laptop with at least 4 GB RAM, core i3 processor, 14'' or above screen, minimum 500 GB HDD, Windows 10 Operating System preloaded, battery, carry case and charger.

## Amendment –II

# Section I: Instructions to Tenderers

## A. Important information at a glance

(The item suffixed by "E" in bracket indicates Eligibility Criteria for a bidder)

### 1. Tender Schedule Details

SCHEDULE	ITEM	QTY	Warranty	Comprehensive Maintenance Contract (CMC)
Schedule-I	Portable Spiro meter with non thermal printing system	27	2 Years	5 Years

## Amendment –III

### Section IV. Schedule of Requirements

#### 1. List of Consignee

Sl. No.	Name of the Consignee	Equipment
		Portable Spiro meter with non thermal printing system
1.	Purulia	1
2.	Paschim Bardhaman	1
3.	Purba Bardhaman	1
4.	Bankura	1
5.	Birbhum	1
6.	Paschim Medinipur	1
7.	Bishnupur	1
8.	Jhargram	1
9.	Rampuhat	1
10.	Darjeeling DH	1
11.	Siliguri DH	1
12.	Kalimpong DH	1
13.	Alipurduar DH	1
14.	Jalpaiguri DH	1
15.	Mathabhanga SDH	1
16.	Gangarampur SDH	1
17.	Islampur SSH	1
18.	Jangipur SSH	1
19.	Krishnananagr DH	1
20.	Imambara DH	1
21.	Kakdwip DH	1
22.	Tamluk DH	1
23.	Contai SDH	1
24.	Onda SSH	1
25.	Raghunathpur SDH	1
26.	MR Bangur DH	1
27.	Chanchal SDH	1
<b>Total</b>		<b>27</b>

Please note:

- The Consignee Receipt Certificate (CRC) will be issued to the Vendor within 72 hours of the delivery at the Consignee address.