



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340308/319
E mail: procurement@wbmsc.gov.in

SUPPLY AND COMMISSIONING OF MEDICAL EQUIPMENTS (SET-III) FOR TRAUMA CARE
FACILITY FOR THE HOSPITALS AND MEDICAL COLLEGES OF THE GOVT. OF WEST
BENGAL

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT- 21/2017

Dated – 26.04.2017

The following amendments have been made in the tender document,

Amendment – V

Read the following wherever mentioned in the bid document:

BI-PAP machine

As

Non Invasive Ventilator

Amendment – VI (Revision of Technical Specification)

The revised technical specifications for the item is given below,

Non Invasive Ventilator

1. Modes of Operation:

- 1.1 Spontaneous
- 1.2 CPAP
- 1.3 **BI-LEVEL**
- 1.4 Variable apnea backup
- 1.5 During changing from one mode to the other, the machine should not be switched off.
- 1.6 Ventilatory circuit including all its parts are to be considered as non-consumables and should be supplied free of cost during warranty and CMC

2. Product Feature:

- 2.1 Simplified standard setting menu
- 2.2 Automatic Leak management
- 2.3 **Automatic /** Adjustable Breath trigger / cycle
- 2.4 Noiseless flow generator of less than 30Db

3. Technical Specification:

3.1 Performance:

- 3.1.1 Operating pressure range: 4 to 25 cm H₂O
- 3.1.2 Maximum single fault pressure: 40 cm H₂O

3.2 Dynamic pressure characteristics:

- 3.2.1 S mode: IPAP: 4 to 25 cm H₂O ; EPAP: 4 to 25 cm H₂O
- 3.2.2 CPAP mode: 4 to 20 cm H₂O

3.3 Sound pressure level:

- 3.3.1 < 30Db (tested in accordance with the requirements of ISO 17510 – 1:2002)

3.4 Display:

- 3.4.1 Leak
- 3.4.2 Pressure
 - 3.4.2.1 IPAP
 - 3.4.2.2 EPAP
- 3.4.3 Respiratory Rate
- 3.4.4 Minute / Tidal Volume

3.5 Power Supply:

- 3.5.1 220 – 240 V AC, 50 – 60 Hz
- 3.5.2 Battery Back-up minimum 1 hour

3.6 Environmental condition:

3.6.1 Operating Temperature – 5 – 40°C

3.6.2 Humidity – 10 – 95%

3.7 Air Filter:

3.7.1 Washable air filter

3.7.2 Filter to be changed whenever needed without disturbance to compressor

3.8 Electromagnetic Compatibility:

3.8.1 Shall meet IEC – 60601 – 1 – 2:2001 (or equivalent BIS) general requirements of safety for electromagnetic compatibility

3.9 Should be **US FDA** approved

3.10 Warranty for 2 years and CMC for 5 years

Amendment – VII

Form 8: STATEMENT OF BREAKUP OF DUTIES AND TAXES

Sl. No.	Particulars	Percentage	Price (In INR)
1	Basic Price of equipment		
2	Customs Duties, if any		
3	Excise Duties, if any		
4	Freight Charges, if any		
5	Insurance, if any		
6	CST, if applicable		
7	VAT, if applicable		
8	GST, if applicable		
9	Others, if any		
BILL VALUE/ INVOICE VALUE, if applicable			
10	Entry tax @ 1% of BILL VALUE/ INVOICE VALUE, if applicable		
Gross Price			

Amendment – VIII

Read Clause 35 (marked in red below) at Page No. 13: Submission and Opening of Bids

iii) BID – C [Bill of Quantity (BOQ), the Statement of Breakup of Duties and Taxes and Cost of Consumables & Spares]

BOQ shall contain the financial quotes in respect of (a) **Gross Price of Equipment (b) Comprehensive Maintenance Contract (CMC) charges.**

- (a) **Gross Price of Equipment (GP):** includes value of goods, accessories & ancillaries and taxes, duties, freight charges, insurance, installation, commissioning, end user training as many number of times as required during the period of warranty and any other charges as applicable. **Entry tax @ 1% should also be considered in the Gross Price.**

AS

iii) BID – C [Bill of Quantity (BOQ), the Statement of Breakup of Duties and Taxes and Cost of Consumables & Spares]

BOQ shall contain the financial quotes in respect of (a) **Basic Price of Equipment (b) Comprehensive Maintenance Contract (CMC) charges.**

Basic Price of Equipment (BP): Basic Price of Equipment(s) includes value of goods, warranty for 2 years, accessories & ancillaries, other charges (e.g. freight charges upto the consignee location, insurance etc), installation , commissioning , end user training as many numbers as may be required during the period of warranty but excluding applicable tax (es).

Amendment – IX

Read Clause 15 (marked in red below) at Page No. 23: Section II. General Conditions of Contract

1. **BLACKLISTING:**

Any bidder / tenderer or its manufacturing agency currently blacklisted by any Government agency in India, is not eligible to participate.

AS

15. BLACKLISTING:

Any manufacturer / its subsidiary which has been blacklisted by any Government Department/ Agency in India during the last five years, would not be eligible to participate in the tender.

Amendment – X

Read Clause Form 1 (marked in red below) at Page No. 57: **Section V. Bidding Forms**

g) We are not currently blacklisted by any Government agency in India.

AS

We have never been blacklisted by any Government Department/ Agency in India during the last 5 years.