



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340308/319
E mail: procurement@wbmsc.gov.in

SUPPLY OF CRITICAL CARE UNIT (CCU) AND HIGH DEPENDENCY UNIT (HDU) EQUIPMENTS
IN THE HOSPITALS AND MEDICAL COLLEGES OF THE GOVERNMENT OF WEST BENGAL.

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT- 147/2018

Dated-24.07.2018

2nd call of tender no. WBMSCL/NIT-99/2018, dated: 06.06.2018(Supply of Ripple Mattress
- Schedule VI)

The following amendment have been made in the tender document,

Amendment –I (Revision of Technical Specification)

The revised technical specifications for the item is given below,

Ripple Mattress

1. Size: Rippled part 6ft X 3ft
2. Alternate rippling effect with air insufflations & sucking effect
3. Air insufflators
4. 2 tubes for alternate air insufflations
5. **Environmental factors**
 - a. The unit shall be capable of being stored continuously in ambient temperature of 10-40° C & relative humidity of 15-90%

- b. Shall meet IEC-60601-1-2:2001 (or equivalent BIS) general requirements of safety for electromagnetic compatibility
- c. Shall be capable of operating continuously in ambient temperature of 10-40° C and relative humidity of 15-90%

6. Power Back-up

- a. Power input: 220-240V/50Hz Single phase or 380-400 V AC50 Hz three phase fitted with appropriated Indian plugs & sockets.

7. Standards, Safety & Training

- a. Electrical safety conforms to standards for electrical safety IEC-60601/IS-13450
- b. Manufacturer should have ISO certification for quality standards
- c. The service provider should have the necessary equipment recommended by the manufacturer to carry out preventive maintenance test as per guidelines provided in the service/maintenance manual
- d. Warranty for 2 years
- e. User manual in English
- f. Service manual in English
- g. List of important spare parts and accessories with their part number and costing
- h. Compliance report to be submitted in a tabulated and point wise manner clearly mentioning the page /para number of original catalogue/data sheet. Any point, if not substantiated with authenticated catalogue/manual, will not be considered.

Note: Initial supply of Ripple Mattress should consist of one unit of air insufflators and one unit of mattress. The remaining 03 (three) units of mattress may be consumed as and when required by the end users.