



## Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited  
Swasthya Sathi  
GN-29, Salt Lake, Sector-V  
Kolkata-700091

Phone No (033) 40340308/319  
E mail: [procurement@wbmsc.gov.in](mailto:procurement@wbmsc.gov.in)

SUPPLY OF NEONATAL CARE EQUIPMENTS FOR FACILITY BASED NEWBORN CARE (FBNC) UNITS IN THE HOSPITALS AND MEDICAL COLLEGES OF THE GOVERNMENT OF WEST BENGAL.

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT- 41/2017

Dated-21.07.2017

The following amendment have been made in the tender document,

### Amendment –II

#### Neonatal Transport Incubator

1. Incubator would be double wall canopy with front and head end access doors with access portholes and tubing access ports.
2. It should have digital displays of air and baby skin temperatures.
3. It should have indicators for mains and battery modes of operation.
4. It should also have indicator for battery power capacity.
5. It must have examination light.
6. It should have AC and 12 V DC connectors,
7. It should have comprehensive alarm system.

8. It should have monitor shelf.
9. It should be supplied with neonatal ventilator with IPPV, IMV and CPAP modes.
10. Adjustable peak pressure (PIP) and PEEP.
11. It should have adjustable Air/ oxygen blender (21-100%).
12. It should have options for wall gas supply or cylinder supply,
13. It should have air compressor.
14. It should have humidification system.
15. It should be supplied with oxygen analyser with digital display of oxygen concentration,
16. It should have facility for suction unit suitable for neonatal use.
17. It should have a trolley to be lightweight on four locking castors with handles.
18. System must be capable of being securely installed in ambulance.
19. Safety and Standard: US FDA & CE ("Conformite Europeene") from European Union notified body having 4 digit identification number Mention the cost of incubator and ventilator separately.
20. Both should have warranty for 2 years and CMC for 5 years.
21. Equipment should be supplied With:
  - a. Ventilator tubings (complete set) - 02 sets.
  - b. Ventilator hose for air and oxygen - 02 in number (2 metres in length).
  - c. Oxygen cylinders compatible with incubator - 02 (should be light weight).
  - d. Air compressor - 01
  - e. Humidifier set - 01
  - f. Internal battery - 01
  - g. Suction unit - 01
  - h. Incubator temperature sensor - 02
22. Mention the cost of the following separately;
  - a. Incubator temperature sensor
  - b. Internal battery
  - c. Ventilator tubing set
  - d. Suction unit

