



## Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited  
Swasthya Sathi  
GN-29, Salt Lake, Sector-V  
Kolkata-700091

Phone No (033) 40340308/319  
E mail: [procurement@wbmsc.gov.in](mailto:procurement@wbmsc.gov.in)

PROCUREMENT OF MEDICAL EQUIPMENT FOR MEDICAL COLLEGE AND HOSPITALS  
(GROUP 3)

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT-217/2018

Dated –12.10.2018

The following amendment has been made in the tender document,

Amendment – 2

### Schedule 14

STEAM STERILIZER :

PARAMETER / FEATURE	THRESHOLD LIMIT / RANGE
TEMPERATURE RANGE IN DEGREE	121
PRESSURE RANGE IN PSI	15-20
INPUT VOLTAGE	220-240V
INPUT LINE FREQUENCY	50HZ

PHASE	<b>SINGLE</b>
INNER CYLLINDER DIAMETER IN mm	400 OR MORE
INNER CYLLINDER DEPTH IN mm	1200 OR MORE
SAFETY VALVE CAN BE SET IN PSI	(10 TO 20) OR MORE
STERILIZER HYDRAULICALLY TESTED IN PSI	40
MAXIMUM POWER CONSUMPTION IN KW	4
<b>MUST HAVE FEATURES</b>	
<b>PARAMETER / FEATURE</b>	<b>YES / NO</b>
STAINLESS STEEL SS304 BOILER	
STAINLESS STEEL OUTER SHELL	
HORIZONTAL FLOOR MOUNTED TYPE	
WING NUTS TIGHTENED STAINLESS STEEL LID	
NEOPRENE RUBBER GASKET	
DIGITAL TEMPERATURE CONTROLLER CUM INDICATOR	
AUTOMATIC PRESSURE CONTROL SWITCH	
LOWER WATER PROTECTION DEVICE	
DIGITAL TIMER	
<b>QUALITY STANDARD AND SAFETY CERTIFICATION</b>	
BIS / EUROPEAN CE / US FDA	