



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340319/308
E mail: procurement@wbmsc.gov.in

SUPPLY OF BILIRUBINOMETER WITH MICROCENTRIFUGE IN THE HOSPITALS AND MEDICAL COLLEGES OF THE GOVERNMENT OF WEST BENGAL

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT-12/2019

Dated –21.01.2019

The following amendments have been made in the tender document,

Amendment – I (Page No. 5)

10. Experience and Technical Capacity (E)

Bilirubinometer with Microcentrifuge	
	Bidder must have supplied Medical equipment worth Rs. 10 Lakh in Govt. Hospitals in India during the last 3 (three) Financial Years (FY)

Amendment – II (Page No. 12)

5	EQUIPMENT	MACHINERIES 2	List of installations & commissioning of offered model (self declaration)
---	-----------	---------------	---

Amendment – III (Page No. 13)

8	Supporting document/ order copy of Medical equipment worth Rs. 10 Lakh supplied in Govt. Hospitals in India during the last 3 (three) Financial Years (FY)		
---	---	--	--

Amendment – IV

(Page No. 34)

REVISED TECHNICAL SPECIFICATIONS

Single Beam Bilirubinometer

1. Specimen required for measurement of bilirubin : whole blood
2. Specimen volume: 60 μ l
3. Type: benchtop
4. Measurement technique: photometric
5. Sensor: silicon photo-diode
6. Optic filter: 455 nm & 575 nm
7. Output range: 4 - 30 mg/ dl or 70 - 500 umol / l
8. Measurement resolution: \pm 0.1 mg/ dl
9. Power: 220 - 240 VAC/ 50 Hz
10. Bilirubin measured: Total
11. Measurement time: 2 to 10 sec
12. Display: LED/OLED
13. Should be supplied with
 - a) Accessories of centrifuge machine: carbon brushes (04) [if applicable], rotor head (01), microhematocrit reader (01)
 - b) Heparinized capillaries-1000 pieces
14. Equipment should be European CE or US FDA approved.
15. It should have warranty for 5 years.
16. External calibration should be done once in three months during warranty periods.
17. Accuracy should be \pm 2%
18. Bilirubinometer should be supplied with inbuilt printer and 5 sets of printer roll. Price of Printer roll should be mentioned separately in Form 10(a). **Date & time should be printed on the test result.**
19. Mention the cost of the following consumables in form 10 (b):
 - a) Carbon brush
 - b) Internal battery if any
 - c) Sensor
 - d) Optic filter (both)
 - e) Lid cover
 - f) Rotor head

MICROCENTRIFUGE

1. It should be a digital microcentrifuge designed for small sample volume.
2. It should have sturdy and durable metal housing.
3. It should have auto-balancing system.
4. Rotor speed should be at least 12000 RPM.
5. It should have safety interlock that can prevent the lid from being opened during operation.
6. Overheating of motor should be prevented.
7. It should have a precision digital timer that can be set from 0 to 15 minutes at 1 min. increment.
8. It should have LED digital display that enables user to monitor the present speed or the remaining time.
9. It should have automatic brake system for rapid deceleration.
10. The lid can be opened automatically/ through switch once the rotor stops spinning.
11. It should accommodate micro or microhematocrit rotors.
12. Should be supplied with lid cover (01)
13. It should comply with quality certificate: ISO9001, 13485

14. It should be CE approved.
15. It should have warranty for 5 years.
Note: The estimated number of Bilirubin test in each Medical College is 30 per day.

Amendment – V

(Page No. 40)

10	List of installations & commissioning of offered model (self declaration)
----	---

Amendment – VI

(Page No. 41)

22	Supporting document/ order copy of Medical equipment worth Rs. 10 Lakh supplied in Govt. Hospitals in India during the last 3 (three) Financial Years (FY)
----	---