



## Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited  
Swasthya Sathi  
GN-29, Salt Lake, Sector-V  
Kolkata-700091

Phone No (033) 40340308/319  
E mail: [procurement@wbmsc.gov.in](mailto:procurement@wbmsc.gov.in)

PROCUREMENT OF MEDICAL EQUIPMENT FOR MEDICAL COLLEGES AND HOSPITALS  
(GROUP 3)

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT-52/2019

Dated –18.06.2019

(2<sup>nd</sup> call of Bid Reference No.: WBMSCL /NIT-217/2018, Dated –12.10.2018)

The following amendment has been made in the tender document,

### Amendment – 1

#### Schedule XIII

STEAM STERILIZER	
PARAMETER / FEATURE	THRESHOLD LIMIT / RANGE
Pressure range in PSI	15-20
Input voltage	220-240V
Input line frequency	50HZ

Phase	Single
Inner cylinder diameter in mm	400 or more
Inner cylinder depth in mm	1200 or more
Safety valve can be set in PSI	(10 to 20) or more
Sterilizer hydraulically tested in PSI	40
<b>MUST HAVE FEATURES</b>	
<b>PARAMETER / FEATURE</b>	<b>YES / NO</b>
Stainless steel SS304 Boiler	
Stainless steel outer Shell	
Horizontal floor mounted type	
Radial locking tightened stainless steel lid	
Neoprene rubber gasket	
Digital temperature controller cum indicator	
Automatic pressure control switch	
Lower water protection device	
Digital timer	
<b>QUALITY STANDARD AND SAFETY CERTIFICATION</b>	
BIS / European CE / US FDA	