



## Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited  
Swasthya Sathi  
GN-29, Salt Lake, Sector-V  
Kolkata-700091

Phone No (033) 40340308/319  
E mail: [procurement@wbmsc.gov.in](mailto:procurement@wbmsc.gov.in)

PROCUREMENT OF MEDICAL EQUIPMENT FOR MEDICAL COLLEGES AND HOSPITALS  
(GROUP 3)

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT-52/2019

Dated –18.06.2019

(2<sup>nd</sup> call of Bid Reference No.: WBMSCL /NIT-217/2018, Dated –12.10.2018)

The following amendment has been made in the tender document,

### Amendment – 1

#### Schedule III

<u>VIDEO BRONCHOSCOPE : (Flex)</u>	
<u>PARAMETER / FEATURE</u>	<u>THRESHOLD LIMIT / RANGE</u>
Outer diameter in mm	6.5 or less
Angulation up side in deg	180 or more
Angulation down side in deg	130 or more
Field of view in deg	120 or more
Depth of field in mm	3 or less
Working Length in mm	600 or more
Instrument channel diameter in mm	2.8 or more

HD LCD monitor size(diagonal) in inch	21 or more
Video Processor output for HDTV monitor (Aspect Ratio)	16:9 or 16:10
Online UPS with battery back up in minute	30 or more
<b>MUST HAVE SPECIFICATIONS</b>	
<b>Parameter / Feature</b>	<b>Yes / No</b>
Computer and Laser Printer with software	
Recording Facility in USB Device	
Should supply Biopsy Forceps: 15 nos.	
Light Source: LED with a facility of NBI/ spectral colour shift/ FICE or 300 W Xenon; additional 2 nos. (i.e. total of 3 nos.) of 300W Xenon lamp should be supplied	
Should supply TBNA needles: 24 nos.	
Should supply Cytology Brush: 20 nos.	
Compact Trolley	
Regular upgradation of software	
<b>Standard &amp; SAFETY Certification</b>	
US FDA or European CE	