



## Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited  
Swasthya Sathi  
GN-29, Salt Lake, Sector-V  
Kolkata-700091

Phone No (033) 40340308/319  
E mail: procurement@wbmsc.gov.in

PROCUREMENT OF MEDICAL EQUIPMENT FOR MEDICAL COLLEGE AND HOSPITALS  
(GROUP 3)

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT-217/2018

Dated –12.10.2018

The following amendment has been made in the tender document,

Amendment – 2

### Schedule II

VIDEO LARYNGOSCOPE : (Rigid)

Parameter / Feature	THRESHOLD LIMIT/Range
Display size (Diagonal) in inch	2 or more
LED life in hours	50,000 or more
EQUIPMENT RUN TIME IN MINUTES WITH FULL BATTERY CHARGE	90 OR MORE
<b>MUST HAVE SPECIFICATION</b>	

Parameter / Feature	Yes / No
Mechanism: integrated camera chip & LED light	
Rechargeable Battery Operated	
Blade (Channeled / non channeled): Macintosh type with size 1, 2, 3 & 4	
ALL BLADES FITS IN A SINGLE HANDLE	
AGREE TO SUPPLY 50 BLADES OF EACH SIZE IN CASE OF DISPOSABLE BLADE/ Reusable 2 sets of each size	
Water Proof	
LIQUID EMERSION DISINFECTION	
<b>Standard &amp; SAFETY Certification</b>	
USFDA & EUROPEAN CE	