

Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340308/319
E mail: procurement@wbmsc.gov.in

PROCUREMENT OF DIFFERENT EQUIPMENTS FOR ESTABLISHMENT OF PHYSICAL MEDICINE & REHABILITATION CENTRE (PMRC) UNDER THE PROJECT IMPLEMENTATION OF HEALTH & FAMILY WELFARE DEPARTMENT.

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT-229/2018

Dated –19.11.2018

The following amendments have been made in the tender document. The changes are highlighted in yellow in the document below,

Amendment – I (Revised Technical Specification)

Item No. 10

Cool Radiofrequency Machine

The equipment should

1. Have touchscreen / LCD or LEDscreen with manual interface,
2. Have 0.50 watt power output,
3. Have thermal, cooled RF & pulsed RF treatment on 1,2,3, or 4 electrodes,
4. Bipolar RF on 1 or 2 pair of electrodes,
5. Measure & display temperature between 0 – 110 degrees
6. Measure impedance 0.3000
7. Have continuous RF output and pulsed RF output functions.
8. Have output test plug to test all outputs form generator before and during the procedure.
9. Have stimulator with pulse duration of 0.1. 0.5, 1,3,3 milliseconds and pulse rate 2-200Hz.

10. Stimulate with 0.01 increments.
11. Gave tuner 0.05-30:00 minutes.
12. E-dose functionality.
13. Capability of epidural electrodes for neuropathic pain.

Should have capability of date export and printing.