

**Notice Inviting e-Tender**  
**West Bengal Medical Services Corporation Limited**  
**Swasthya Bhawan Complex**  
**GN-29, Salt Lake, Sector-V**  
**Kolkata-700091**

Phone No (033) 23577070/40340307/40340308  
E mail: [deo.wbmisc@gmail.com](mailto:deo.wbmisc@gmail.com), [md.wbmisc@gmail.com](mailto:md.wbmisc@gmail.com)

**SUPPLY OF BLOOD-BANK EQUIPMENTS IN THE HOSPITALS AND MEDICAL COLLEGES OF THE GOVERNMENT OF WEST BENGAL.**  
(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT- 48 /2014

Dated- 02.08.2014

**The following amendment have been made in the tender document,**

**Amendment - IX**

**UPDATED & REVISED TECHNICAL SPECIFICATION**

**SCHEDULE - IV**

**Donor Station (Donor Bed with Couch System)**

- |                                  |  |
|----------------------------------|--|
| (a) Power source                 | : 230 ± 10% VAC, 50 Hz, single phase   |
| (b) Lifting capacity (approx.)   | : 150 kg (Donor Weight).   |
| (c) Interface                    | : For connecting Blood Collection Monitor  |
| (d) Classification               | : Class I, (similar to Type B applied part) protection against electric shock  |
| (e) Power Consumption            | : Operating 100W (max.) /Standby 5W max.   |
| (f) Movement / Position          | : Any intermediate position between seating position / donation position and vasovagal attack position. Head end side & foot end side should be adjustable |
| (g) Movement actuation           | : DC motor and torque multiplier   |
| (h) Initiation of motor movement | : By operating the rocker switch / remote control  |
| (i) Arm rests                    | : Both swivel able and height adjustable   |
| (j) Upholstery                   | : Soft moulded washable (easily cleanable) upholstery of 2.5" thickness  |

(k) Manoeuverability	: Provided with lockable castor wheels
(l) Accessories	: BCM tray, BP apparatus tray, tray for Accessories and telescopic IV stand.
(m) Weight	: 75 – 90 kg
(n) Length of fixed seat	: 100 cm +/- 5 cm
(o) Length of swivel seat	: 90 cm +/- 5 cm
(p) Width	: 67 cm +/-5 cm
(q) Length of arm rest	: 60 cm +/- 5 cm
(r) Width of arm rest	: 15 cm +/- 2 cm
(s) Height of leg rest	: 80 cm +/- 5 cm
(t) Height of Seat	: 51 cm +/- 5 cm
(u) Vasovagal attack position	: 66 cm +/- 5 cm
(v) Max. height of back rest	: 110 cm +/- 5 cm
(w) Fuse	: 250 V, 1A, 5 * 20mm type M, SB
(x) Standards & approvals	: CE / FDA / BIS or any other indian or internationally approved safety standard certification
(y) Manufacturing Standard	: ISO 9001:2008 or any other indian or internationally approved standard certification, EN ISO 13485:2012 (for electrical & bio safety hazard)
(z) Warning	: Replace fuses with specified type and rating