



**COVID-19  
URGENT**

## Notice Inviting e-Tender

**West Bengal Medical Services Corporation Limited  
Swasthya Sathi  
GN-29, Salt Lake, Sector-V  
Kolkata-700091**

Phone No (033) 40340307 / 40340308

E mail: [procurement@wbmsc.gov.in](mailto:procurement@wbmsc.gov.in)

**Procurement of Finger Tip Pulse Oximeter for COVID-19 purpose**  
(Submission of Bid through *online*)

Bid Reference No.: WBMSCL / NIT-100 / 2020

Dated - 14.07.2020

The following amendments have been made in the tender document,

### **Amendment – II**

(Page No. 2)

- i) **Delivery Timeline:** Total supply to be completed within **14 days** in **2 tranches** as per following schedule

1 <sup>st</sup> tranche	Within <b>7 days</b> after issuance of Award of Contract (AOC)	<b>5000 units</b>
2 <sup>nd</sup> tranche	Within <b>14 days</b> after issuance of Award of Contract (AOC)	<b>5000 units</b>

### **Amendment – III**

(Page No. 3)

#### ANNEXURE A

#### **Specification of Fingertip Pulse Oximeter**

1. Measurement Range: SpO2 Range: 35 to 100% and Pulse Rate: 30 to 250bpm
2. Accuracy Range: SpO2:  $\pm 2\%$  (90-100%),  $\pm 3\%$  (70-90%), Unspecified: 70%, Pulse Rate:  $\pm 1\%$  or  $\pm 3$ bpm
3. **Should have Manual/ Auto-Rotate Screen.**
4. Color-changing visual alerts.
5. Low Battery Indication.
6. Energy Efficient.
7. Battery backup 1.5hours continuously or more.
8. High coverage Finger Pad.
9. Replacement Warranty: 1 year
10. Compact & Light weight Design.
11. **Should be European CE/ BIS approved.**