

**Notice Inviting e-Tender**  
**West Bengal Medical Services Corporation Limited**  
**Swasthya Bhawan Complex**  
**GN-29, Salt Lake, Sector-V**  
**Kolkata-700091**

Phone No (033) 23577070/40340307/40340308

E mail: [procurement@wbmsc.gov.in](mailto:procurement@wbmsc.gov.in), [md@wbmsc.gov.in](mailto:md@wbmsc.gov.in)

**SUPPLY OF BLOOD-BANK EQUIPMENTS IN THE HOSPITALS AND MEDICAL COLLEGES OF THE GOVERNMENT OF WEST BENGAL.**

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT-14/2015

Dated- 25.03.2015

**The following amendment have been made in the tender document,**

**Amendment - II**

**REVISED TECHNICAL SPECIFICATION**

**Schedule VIII**

**INCUBATOR WITH THERMOSTATIC CONTROL (75 LTR.)**

- a) Control: Equipped with Microprocessor controlled with a large digital display.
- b) Work Chamber Volume: minimum 75 Litres.
- c) Working temp.: From room temp. +5 °C to 75 °C
- d) Temp. deviation with respect to time:  $\pm 0.2$  °C.
- e) Display: Digital display located on front panel of the equipment.
- f) Easy to clean: An interchangeable silicon door gasket & removal of shelf supports make thorough cleaning and decontamination of the working chamber easy & safe
- g) Inner casing: Should be made of corrosion resistant stainless steel.
- h) Shelves **(adjustable)**: Perforated shelves made of corrosion resistant stainless steel. Standard supply 3 nos. , inter shelves gap – 6" approx.
- i) Heat conduction: Gravity convection or air circulation.
- j) Heater on rear / side / bottom of the equipment.
- k) Power requirement: 230 V, single phase, 50 Hz.
- l) After sales service: Prompt & efficient after sales service should be available from Kolkata.
- m) Product certification: CE / US FDA / BIS certified
- n) Quality Certification: ISO certificate