



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Bhawan Complex
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340307 / 40340308

E mail: procurement@wbmsc.gov.in

PROCUREMENT OF MEDICAL EQUIPMENT FOR FIVE NEW MEDICAL COLLEGE OF THE
GOVT. OF WEST BENGAL

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT- 213/2018

Dated -11.10.2018

Amendment – VI

Technical Specifications

Amendment VI: The specification of the items put out in the tender is detailed below

Sl. No.	Name of the discipline	Name of the Equipments (Item)	Specifications																								
	Package 2																										
30	Physiology	Tuning fork	128 Hz should be made up of stainless steel with frequency marked 256 Hz should be made up of stainless steel with frequency marked 512 Hz should be made up of stainless steel with frequency marked																								
	Package 3																										
33	Biochemistry	Fixed volume pipettes	<ul style="list-style-type: none"> • Should be non autoclavable • Indicative make: Sartorius/Pipetteman/Eppendorf/Gilson/Biohit/Thermofisher • Set of 5 x 5 micropipettes Set of Micropipettes consists of the following:																								
			<table border="1"> <thead> <tr> <th>vol. range</th> <th>Accuracy (±%)</th> <th>CV (±%)</th> <th>No. required</th> </tr> </thead> <tbody> <tr> <td>1000ul</td> <td>0.5%</td> <td>0.2</td> <td>5</td> </tr> <tr> <td>500ul</td> <td>0.5%</td> <td>0.2</td> <td>5</td> </tr> <tr> <td>100ul</td> <td>0.5%</td> <td>0.2</td> <td>5</td> </tr> <tr> <td>10ul</td> <td>1%</td> <td>0.5</td> <td>5</td> </tr> <tr> <td>5ul</td> <td>2%</td> <td>1</td> <td>5</td> </tr> </tbody> </table>	vol. range	Accuracy (±%)	CV (±%)	No. required	1000ul	0.5%	0.2	5	500ul	0.5%	0.2	5	100ul	0.5%	0.2	5	10ul	1%	0.5	5	5ul	2%	1	5
			vol. range	Accuracy (±%)	CV (±%)	No. required																					
			1000ul	0.5%	0.2	5																					
			500ul	0.5%	0.2	5																					
			100ul	0.5%	0.2	5																					
10ul	1%	0.5	5																								
5ul	2%	1	5																								