



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340308/319
E mail: procurement@wbmsc.gov.in

SUPPLY OF AUTOMATED CELL COUNTER IN THE HOSPITALS AND MEDICAL COLLEGES OF THE GOVERNMENT OF WEST BENGAL.

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT-51/2019

Dated –14.06.2019

The following amendments have been made in the tender document,

Amendment –I (Revision of Technical Specification)

The revised technical specifications for the item is given below,

Automated Cell Counter

1. The machine should be Fully Automated Haematology Analyzer providing at least 19 parameters including 3-part differential, with user definable settings to have both RDW-CV & RDW-SD.
2. Cost per test, onboard stability & test pack size to be included in Form 8 & Form 11.
3. The system should be capable of processing samples at a speed of at least 60 samples per hour in whole blood mode. Pediatric mode should be there for analyzing low volume samples with reportable results for differential parameters along with total counts.
4. The system should use the proven and approved “Volumetric Metering” system of cell counting for WBC’s, RBC’s and PLT’s for high precision of the results & stability of the calibration. WBC differential parameters such as Granulocytes / Neut, Lym & Mixed Cells in absolute count will be preferred.

5. The system should have a large LCD / LED display to have a review of all results along with the three histograms of WBC, RBC & PLT on the screen with automatic floating discriminator.
6. The system should be a two reagent system other than a cleaner with non-cyanide based reagent for Hb estimation, using high intensity LED and not lamp.
7. The system should have world reference "Electrical Impedance" method of cell counting for the reliability of the results, with an integrated temperature sensor for monitoring & compensating for shifts in room temperature. The system should have inbuilt probe wiping after every sample.
8. 4 sets of tri-level control to be supplied in a staggered manner over a period of 6 months.
9. External Roller Mixture with capacity of at least 6-9 vials to be supplied.
10. Equipments should be supplied (staggered) with a free reagent for testing 1000 samples in a span of 6 months.
11. The system should have a low cost per test (considering work load of 400 samples per month).
12. Should be supplied with 1KVA online UPS (Libert / APC) having 30 minutes back up.
13. All reagents required should be available locally from the company or its authorized distributor within 7 days of placement of order.
14. Warranty: 2 years; CMC: 5 years.
15. Product certification: CE (Declaration of Conformity on compliance with In Vitro Diagnostic Device Directive IVD – 98/79/EC) or US FDA Certified.

Note: Cost for consumable required for testing and obtaining 24000 reportable test results likely to be undertaken for 5 years will be considered during financial bid evaluation. The cost for the consumable will remain unchanged during 5 years.

Amendment –II

Form 8(a) under Section V. Bidding Forms has been modified. The amended format is given below:

Form 8(a): Details of Reagents along with calculation of cost per reportable test result.

Equipment:								
Sl. No.	Item	Pack Size	Cost of pack excluding GST	Volume required in one test	Cost for one test excluding GST (a)	Volume required for shut down and start up once per day	Cost for one test excluding GST (b)	Total cost (c) = (a) + (b)
1	Diluent							
2	Lyse							
3	Cleaner							
4	Other, if any							
5	Other, if any							
Total [(c) = (a) + (b)] [Test per reportable test result]								

The above calculation is excluding cost of calibration.

However, one set of Calibrator and one set of tri-level control should be supplied free of cost with the equipment on receipt of confirmation from the consignee about installation and commissioning of the equipment.

NOTE: The Tenderers should furnish price of all the Consumables of the offered model in the above format. If any Tenderer do not quote price of any of the consumables required for functioning of the equipment, it will be presumed that those items will be supplied FREE OF COST by the Tenderer for 5 years.