



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340308/319
E mail: procurement@wbmsc.gov.in

SUPPLY AND COMMISSIONING OF MEDICAL EQUIPMENTS (SET-III) FOR TRAUMA CARE FACILITY FOR THE HOSPITALS AND MEDICAL COLLEGES OF THE GOVT. OF WEST BENGAL

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT- 66/2017

Dated – 23.09.2017

2nd call of Bid reference no. WBMSCL/NIT-21/2017, dated: 26.04.2017 of Schedule –I, II, III, IV & V

SCHEDULE – V

The following amendments have been made in the tender document,

Amendment – IV

Patient Warming System

1. External warming of the patient while providing full, unrestricted patient access
2. Fluid outlets should minimize pooling of fluids on the surface of the blanket and water proof tear proof blanket
3. Allows for flexible patient positioning
4. Accommodates the use of a drawsheet (where applicable)
5. Soft, radiolucent, latex-free materials
6. Resealable hose ports

7. OPTIONS OF FOLLOWING TYPES OF BLANKETS SHOULD BE AVAILABLE:

8. Intraoperative blankets: Upper Body Blanket / Lower Body Blanket – 100 unit
Each

9. Paediatric blankets: Postoperative Pediatric Full Body - 100 units

10. Speciality blankets: Full Body Surgical Blanket - 100 units

11. ICU blankets: Full Body Blanket - 100 units

12. Optional: (A) Intraoperative blankets: XL Upper Body Blanket / Torso Blanket / Dual Port Torso Blanket; (B) Underbody blankets: Pediatric Underbody Blanket / Large Pediatric Underbody Blanket / Adult underbody Blanket / Spinal Underbody Blanket / Lithotomy Underbody Blanket / Full Access Underbody / Presealed Sterile Full Access Underbody Blanket; (C) Paediatric blankets: Small Lower Body Blanket / Pediatric Long Blanket / Pediatric Underbody Blanket / Large Pediatric Underbody Blanket ; (D) Speciality blankets: Preoperative & Outpatient Care Blanket / Surgical Access Blanket ; (E) ICU blankets: Chest Access Blanket / Postoperative Pediatric Full Body Blanket / Multi-Access Blanket.

13. TEMPERATURE MANAGEMENT UNITS:

14. Forced-air warming

15. Small, lightweight and easy to use

16. Should provide clinically effective patient warming at safe air delivery temperatures

17. Should meet regulatory standards for leakage current

18. Licenses of national and international (USA / Europe) regulatory bodies should be available

19. System should have safety certificate from a competent authority CE / US FDA

20. Filter life should be 1000 hrs.