

Notice Inviting e-Tender
West Bengal Medical Services Corporation Limited
Swasthya Bhawan Complex
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 23577070/40340307/40340308

E mail: deo.wbmisc@gmail.com, md.wbmisc@gmail.com

SUPPLY OF BLOOD-BANK EQUIPMENTS IN THE HOSPITALS AND MEDICAL COLLEGES OF THE GOVERNMENT OF WEST BENGAL.

(Submission of Bid through *online*)

Bid Reference No.: WBMISCL/NIT- 48 /2014

Dated- 02.08.2014

The following amendment have been made in the tender document,

Amendment -X

UPDATED & REVISED TECHNICAL SPECIFICATION

SCHEDULE - X

Centrifuge (Table Top)

- a) Capacity : 16 holes x15ml
- b) Built in time and speed regulator (digital controlled) with suitable speed indication (digital display) and lid lock system
- c) RPM : 500 to 6000 rpm
- d) Power Supply : 220/240 volts, single phase, 50 Hz
- e) A line voltage corrector of suitable rating should form part of the configuration as per IS:9815/89
- f) Technical Literature: The firm shall submit printed illustrated technical literature/ leaflet indicating the model number. If quoted model is a modified version of their any standard product that also be indicated in the offer
- g) Standards & approvals : CE / FDA / BIS or any other indian or internationally approved safety standard certification

h) Manufacturing Standard : ISO 9001:2008 or any other indian or internationally approved standard certification, EN ISO 13485:2012 (for electrical & biosafety hazard)