

Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited Swasthya Sathi GN-29, Salt Lake, Sector-V Kolkata-700091

Phone No (033) 40340308/412 E mail: procurement@wbmsc.gov.in

Procurement of equipment for CCU/HDU and other for COVID 19 Hospitals of the State of West Bengal.

(Submission of Bid through online)

Bid Reference No.: WBMSCL /NIT-205/2020 Dated -13.10.2020

The following amendment have been made in the tender document,

<u>Amendment –I (Revision of Technical Specification)</u>

The revised technical specifications for the item is given below,

X-RAY Machine

PARAMETER/FEATURE	THRESHOLD LIMIT/ RANGE
Generator	
Frequency in KHZ	100 or more
Output in KW	4 or more
KV range	40 to100
mA range	upto 100
mAs	200 or more
X-ray tube	
Anode focal spot size in mm	2.8×2.8 or less
Operating voltage range in Volt	190-240
Current in Amp	15
MUST HAVE SPECIFICATION	

PARAMETER/ FEATURE	YES/ NO
Operational Requirements	
Compact, lightweight, easily transportable,	
mobile unit	
Effective Braking system for parking and	
transport	
Spring balanced or counter balanced tube	
stand	
Tube head rotate in all direction	
High frequency (not biphasic)	
Digital display of KV& mAs	
Exposure with remote controlled hand	
switch (Wired or Remote)	
Generator	
Microprocessor controlled	
X Ray tube	
Stationary Anode	
Manual collimator	
LED for visualization of Radiation field	
Auto shut off facility for Collimator	
Run on Single phase	
Accessories	
Lead Jacket-04 NOS.	
Lead Gogles-01 NOS.	
Thyroid Guard-02 NOS.	
Lead Hanger-01 NOS.	
Quality Standards & Safety Certification	
ISO	
AERB type approval	
European CE/ BIS	