

Notice Inviting e-Tender
West Bengal Medical Services Corporation Limited
Swasthya Bhawan Complex
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 23577070/40340307/40340308

E mail: deo.wbmisc@gmail.com, md.wbmisc@gmail.com

SUPPLY OF BLOOD-BANK EQUIPMENTS IN THE HOSPITALS AND MEDICAL COLLEGES OF THE GOVERNMENT OF WEST BENGAL.

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT- 48 /2014

Dated- 02.08.2014

The following amendment have been made in the tender document,

Amendment - I

REVISED TECHNICAL SPECIFICATION

SCHEDULE - VII

Elisa Reader

- (a) Photometer 8 - 12 measurement channel
- (b) Wave length range 400-750 nm
- (c) Filter: 405, 450,492 & 620 nm with optional additional filters.
- (d) Absorbance range 0 to 4.0 OD
- (e) Single wave length 6 seconds or better reading
- (f) Scan time dual wave length 8 seconds or better
- (g) Multi level volume measurement
- (h) On board shaking with time & speed
- (i) Halogen lamp / Xeon / LED
- (j) Lamp shelf life: At least 1 million reading

(k) System should be CE / FDA / BIS/ISO or any other internationally approved safety standard certification

(l) Compatible UPS – 30 minutes backup

Programming:

(a) Computer controlled with windows based software, provided with compatible external laser jet printer

(b) Multi blank programming facility, various types of cut off program

(c) 8 x 12 matrix print facility

Elisa Washer

a) Cross wise aspiration, over flow washing, bottom washing

b) Automatic manifold detection

c) 8 or 12 channel manifold

d) Manifold user changeable, plate carrier auto cleavable and also can be sterilized

e) Equipment should be, capable of using any bottle or container

f) Volume of wash liquid dispensed 50 to 350 μ l with 50 μ l increment

g) Suitable for UV & flat bottom micro plate etc.

h) Residual volume per well < 2 μ l

i) System should be CE / FDA / BIS/ ISO or any other internationally approved safety standard certificate