



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340308/319
E mail: procurement@wbmsc.gov.in

SUPPLY OF NEONATAL CARE EQUIPMENTS FOR FACILITY BASED NEWBORN CARE (FBNC) UNITS IN THE HOSPITALS AND MEDICAL COLLEGES OF THE GOVERNMENT OF WEST BENGAL.

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT- 104/2017

Dated-29.11.2017

3rd call of Bid reference no. WBMSCL/NIT-41/2017, dated: 21.07.2017 of Schedule – V and VIII

The following amendments have been made in the tender document,

Amendment – I

Read the following wherever mentioned in the bid document:

Air Oxygen Blender (Mechanical)

As

Heated Humidified High Flow Nasal Cannula (HHHFNC) System for Newborns

Amendment – II (Revision of Technical Specification)

The revised technical specifications for the item is given below,

Heated Humidified High Flow Nasal Cannula (HHHFNC) System for Newborns

Blender:

1. Blender should deliver oxygen concentration range (21% to 100%)
2. Flow rate 1- 15 litre/min
3. Accuracy $\pm 3\%$

Humidifier:

4. Servo controlled humidifier base with digital temperature display with audio visual alarms for conditions like high & low temperature, humidity & disconnection and / more. Humidifier should be US FDA and CE ("Conformite Europeene") from European Union notified body having 4 digit identification number approved product.

Circuit:

5. Heated Breathing Circuit with heater wire technology to provide proper humidification, at least 5 reusable sets and at least 20 disposables with each machine.
6. Should have facility for high flow nasal Cannula. 15 nos. of Cannula (Small – 5 nos., Extra Small – 5 and Large - 5) should be supplied with machine.

Accessories:

1. Air and oxygen hose and coupler
2. Bracket pole mount
3. Bracket wall mount
4. Trolley

Standards:

1. Should have the ISO certification and the copy of the same should be enclosed along with the technical bid.
2. Should comply with CE ("Conformite Europeene") from European Union notified body having 4 digit identification number / BIS standards

Documentation:

User/Technical/Maintenance manual to be supplied in English.

Warranty & CMC

The machine and all other accessories should be under warranty for 2 years and followed by 5 years CMC.