



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited

Swasthya Sathi

GN-29, Salt Lake, Sector-V

Kolkata-700091

Phone No (033) 40340307/320

E mail: procurement@wbmsc.gov.in

Supply and Commissioning of Neuro Surgical Head Fixator Systems for BIN -Annex of IPGME&R-SSKM Hospital

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT-069/2023

Dated-17.02.2023

Amendment-I

Revised Technical Specification

Neuro Surgical Head Fixator Systems

- Should be constructed of corrosion and rust resistant durable metal.
- Disposable attachments and consumables such as fixation pins should be readily available and vendor should provide the items from time to time as orders are placed.
- Adult and paediatric type pins (guarded) should be available.
- Compression gauge should be provided with clear display of compression force/pressure in SI units.
- Each unit should contain one 3 pin head holder and one table attachment.
- Units should be compatible with Maquet AlphaClassic or AlphaMax OT table (as installed in BIN OT).
- Units should be compatible of various Neurosurgical positions, including advanced and complex positions such as sitting position, park bench position, Fukushima position etc.
- Units should be compatible with attachments of Neuronavigation systems such as Brain labor Medtronic Stealth station S8 (Universal Mayfield type screw adaptor).
- Equipment should be European CE (4 Digit notified Body) and USFDA certified.