



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340308/319
E mail: procurement@wbmsc.gov.in

SUPPLY OF NEONATAL CARE EQUIPMENTS FOR FACILITY BASED NEWBORN CARE (FBNC) UNITS IN THE HOSPITALS AND MEDICAL COLLEGES OF THE GOVERNMENT OF WEST BENGAL.

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT- 41/2017

Dated-21.07.2017

The following amendment have been made in the tender document,

Amendment –II

Air Oxygen Blender **(Mechanical)**

1. Blender should deliver oxygen concentration range (21% to 100%)
2. Flow rate 1- 15 litre/min
3. Accuracy $\pm 3\%$
4. Size approx. (9 X 14 X 12 cm)²
5. Weight approx. (1.5 kg)²

Accessories:

1. Air and oxygen hose and coupler
2. Bracket pole mount
3. Bracket wall mount

Standards:

1. Should have the ISO certification and the copy of the same should be enclosed along with the technical bid.
2. Should comply with CE ("Conformite Europeene") from European Union notified body having 4 digit identification number / BIS standards

Documentation

User/Technical/Maintenance manual to be supplied in English.

Warranty & CMC

The machine and all other accessories should be under warranty for 2 years and followed by 5 years CMC.

Amendment –III

Form 9(a): PRICES FOR CONSUMABLES (to be uploaded for each schedule)

Sl. No	Items	Basic Price of 1(one) number in INR excluding GST
1		
2		
3		
4		

NOTE: The Tenderers should furnish the price of all the Consumables of the offered model in a separate sheet. If any Tenderer do not submit the price of any of the consumables, it will be presumed that those Consumables shall be supplied FREE OF COST by the Tenderer during Warranty & CMC period.

