



Notice Inviting e-Tender
West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340308/319

E mail: procurement@wbmsc.gov.in / mp@wbmsc.gov.in

**Procurement of Medical Equipment for Medical College and Hospitals
(GROUP-5)**

(Submission of Bid through *online*)

Bid Reference No.: WBMSC/NIT- 218 /2018

Dated-12.10.2018

The following amendments have been made in the tender document,

AMENDMENT -II

SCHEDULE I

TECHNICAL SPECIFICATION

SLIT LAMP

Parameter / Feature	Threshold Limit / Range
Magnification selection in step drum rotation	5
Magnification ratio	6X,10X,16X,25X & 40X or more
Eyepieces	12.5 X
Slit width in mm	0-14
Slit length in mm	1-14
Slit angle in degree	0-180

MUST HAVE SPECIFICATIONS

Parameter / Feature	Yes / No
Galileo convergence	
Blue, red free, amber UV cut, IR cut	
LED light source	
Applanation tonometer (hanging type with standard features)	
Integrated high resolution camera with Computer System	
Motorised Table	
Electrical Safety and Quality Certificaton	
European CE & US FDA approved	