



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340308/319
E mail: procurement@wbmsc.gov.in

PROCUREMENT OF MEDICAL EQUIPMENT FOR MEDICAL COLLEGES AND HOSPITALS
(GROUP 3)

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT-52/2019

Dated –18.06.2019

(2nd call of Bid Reference No.: WBMSCL /NIT-217/2018, Dated –12.10.2018)

The following amendment has been made in the tender document,

Amendment – 1

Schedule VI

BI PAP MACHINE	
PARAMETER/ FEATURE	THRESHOLD LIMIT/ RANGE
Inlet port O ₂ flow in L/min	Upto 30
IPAP pressure range in cm of H ₂ O	4 to 30 or more
EPAP pressure range in cm of H ₂ O	4 to 20 or more

Ti-control setting in sec (max)	3 or more
Ti-control setting in sec (min)	0.1
Respiratory rate in BPM	5 to 40 or more
Rise time in m.sec	150-400 or more
Weight in kg	3 or less
MUST HAVE SPECIFICATION	
PARAMETER/ FEATURE	YES/ NO
Mode: ST (spontaneous/timed), Pressure Assisted Control(PAC), CPAP, Spontaneous, Timed, Volume Assured Pressure Support	
Leak compensation and synchronization	
Color screen for real time monitoring	
Alarms for patients safety	
Use in invasive tracheostomy patients	
Inbuilt battery backup for atleast 2 hours	
Humidifier and air filter	
Patient reminder for filter and mask replacements	
European CE and US FDA Certified	