



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340308/319
E mail: procurement@wbmsc.gov.in

PROCUREMENT OF OPERATING HYSTEROSCOPE WITH ACCESSORIES & ENDOMAT AT
WALSH SDH SREERAMPORE

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT-202/2020

Dated –13.10.2020

The following amendments have been made in the tender document,

Amendment – I (Page No. 14)

The cost for Year wise CMC charges after completion of 2 years warranty in percentage of the sum of quoted price of the equipment in the BOQ to be paid to the selected bidder(s) in the following manner:

Operating Hysteroscope with accessories and Endomat				
CMC charges in percentage(%) for Year3	CMC charges in percentage(%) for Year4	CMC charges in percentage(%) for Year5	CMC charges in percentage(%) for Year6	CMC charges in percentage(%) for Year7
4	4.25	4.5	4.75	5.0

Amendment – II

(Page No. 34-35)

Technical Specification for diagnostic & operative Hysteroscopy set

Sl. No.	Description	Qty	Covered under warranty & CMC
1	Hysteroscopy Diagnostic Sheath for 2.9 mm telescope	01 Nos.	Y (2 years)
2	Full HD Telescope 2.9 mm 30 Degree with Rod Lenses, Optical system, Autoclavable with incorporated fiber optic light transmission length minimum 200 mm; Should have European CE	01 Nos.	Manufacturing warranty for 1 year
3	Operative sheath(Inner &Outer) for 2.9 mm forward oblique telescope 30 degree with 5 Fr instrument Channel Auto lock system with Discreet inflow & outflow channels	01 Nos.	Y (2 years)
4	Semi Rigid Scissor Single Action 5Fr	01 Nos.	Y (2 years)
5	Semi Rigid Alligator / Grasping Forcep 5 Fr	01 Nos.	Y (2 years)
6	Semi Rigid Biopsy Forcep 5 Fr	01 Nos.	Y (2 years)
7	Bipolar L Hook 5 Fr	01 Nos.	Y (2 years)
8	Continuous Flow Inner outer Sheath with Obturator For 2.9 mm Telescope for TCRE	01 Nos.	Y (2 years)
9	Mono as well as Bipolar working Element	01 Nos.	Y (2 years)
10	Cord For Working Element	01 Nos.	Y (2 years)
11	Bipolar Cutting Loop	01 Nos.	Y (2 years)
12	Bipolar Ball Electrode	01 Nos.	Y (2 years)
13	Collins knife	01 Nos.	Y (2 years)
14	Tubal Recanalization Guide wire	01 Nos.	Y (2 years)
15	Ellik Evacuator	01 Nos.	Y (2 years)
16	Full HD Telescope 4 mm 30 Degree with Rod Lenses, Optical system, Autoclavable with incorporated fiber optic light transmission length minimum 200mm; Should have European CE	01 Nos.	Manufacturing warranty for 1 year
17	Diagnostic Sheaths	01 Nos.	Y (2 years)
18	Operative sheath(Inner &Outer) for 4 mm forward oblique telescope 30 degree with 5Fr to 7 Fr instrument Channel Auto lock system with Discreet inflow & outflow channels Operative Sheath Continuous flow	01 Nos.	Y (2 years)
19	Mono polar Bug Bee Electrode	01 Nos.	Y (2 years)
20	Semi Rigid Scissor Single Action 5Fr to 7 Fr	01 Nos.	Y (2 years)
21	Semi Rigid Alligator/Grasper 5Fr to 7 Fr	01 Nos.	Y (2 years)
22	Semi Rigid Biopsy Forcep 5Fr to 7 Fr	01 Nos.	Y (2 years)
23	Continuous Flow Inner Outer Sheath For 4 mm Telescope for TCRE	01 Nos.	Y (2 years)
24	Mono as well as bipolar Working Element. For 4 mm Telescope	01 Nos.	Y (2 years)
25	Loop Straight, 90 Degree	01 Nos.	Y (2 years)
26	Collins knife	01 Nos.	Y (2 years)
27	Endomat/ Hysteropump <ul style="list-style-type: none"> ➤ Flow & Pressure controlled Irrigation System ➤ Liquid Pressure Monitoring Technology ➤ Disposable Pressure transducer to avoid Patient Fluid Contamination ➤ Real Time Pressure & Flow Monitoring of Intra-Uterine 	01 Nos.	Y (2 years)

Sl. No.	Description	Qty	Covered under warranty & CMC
	<p>Cavity I specific surgical site</p> <ul style="list-style-type: none"> ➤ Built-in Low Noise irrigation pump ➤ Surgery Specific Custom Modes : Hysteroscopy & versatile ➤ Preferably indicator for amount of consumed fluid by the patient ➤ Preferably for Saline & Glycine Mode settings Alarm indication for consumed fluid after safety threshold ➤ Priming / Flush for removing air bubble from tubing at initialization ➤ Touch Screen LCD Interface <p>Pressure Range: 10 to 400 mmHg \pm 10% Flow Range: Irrigation : 100mL to 2000mL \pm 10% Pressure Setting: Irrigation : 10mmHg 1 step Flow Setting: Irrigation : 100mL 1 step Operating Modes: Only Irrigation Mode Filter: Disposable Cartridge Filter (0.2 micron) Tubing : Silicon Tubing Tube Specifications: Silicon tubing for Irrigation and Suction of 3Mtr Touch Control Panel: 5" LCD Screen with brightness control To be supplied with reusable Irrigation Tubing Set</p>		