



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340308/319
E mail: procurement@wbmsc.gov.in

PROCUREMENT OF MEDICAL EQUIPMENT FOR MEDICAL COLLEGE AND HOSPITALS
(GROUP 3)

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT-217/2018

Dated –12.10.2018

The following amendment has been made in the tender document,

Amendment – 2

Schedule XIII

ANALYTICAL BALANCE :

PARAMETER / FEATURE	THRESHOLD LIMIT / RANGE
MEASUREMENT RANGE	0.1 mg-200gm
PAN SIZE MINIMUM IN mm	80 or more
REPEATABILITY IN mg	0.2
LINE VOLTAGE	100 - 240 V, 50-60HZ
READABILITY IN mg	0.1
MUST HAVE FEATURES	

PARAMETER / FEATURE	YES / NO
GLASS CASING	
DIGITAL DISPLAY	
AUTOMATED TARING SYSTEM	
CHEMICAL RESISTANCE	
INTERNAL CALIBRATION	
QUALITY STANDARD AND SAFETY CERTIFICATION	
BIS / EUROPEAN CE	
LEGAL METROLOGY	