



## Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited  
Swasthya Sathi  
GN-29, Salt Lake, Sector-V  
Kolkata-700091

Phone No (033) 40340308/319  
E mail: [procurement@wbmsc.gov.in](mailto:procurement@wbmsc.gov.in)

PROCUREMENT OF MEDICAL EQUIPMENT FOR MEDICAL COLLEGES AND HOSPITALS  
(GROUP 3)

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT-52/2019

Dated –18.06.2019

(2<sup>nd</sup> call of Bid Reference No.: WBMSCL /NIT-217/2018, Dated –12.10.2018)

The following amendment has been made in the tender document,

### Amendment – 1

#### Schedule II

VIDEO LARYNGOSCOPE : (Rigid)	
Parameter / Feature	THRESHOLD LIMIT/Range
Display size (Diagonal) in inch	2 or more
LED life in hours	50,000 or more
Equipment run time in minutes with full battery charge	80 or more
MUST HAVE SPECIFICATION	
Parameter / Feature	Yes / No
Mechanism: integrated camera chip & LED light	
Rechargeable battery operated (preferable)	
Blade (channeled / non channeled): macintosh type with size 0,1, 2, 3 & 4 (any four)	
All blades fits in a single handle	
Agree to supply reusable 1 set of each size	
Water proof	
Liquid emersion disinfection	
Standard & SAFETY Certification	
US FDA & European CE	