

Notice Inviting e-Tender

**West Bengal Medical Services Corporation Limited
Swasthya Bhawan Complex
GN-29, Salt Lake, Sector-V
Kolkata-700091**

Phone No (033) 2357-7070/40340307/40340308
E mail: deo.wbmisc@gmail.com, md.wbmisc@gmail.com

**SUPPLY OF AUTOMATED BLOOD GAS ANALYSER, SEMI AUTOMATED
BIOCHEMISTRY ANALYSER AND ELECTROLYTE ANALYSER IN THE HOSPITALS
AND MEDICAL COLLEGES OF THE GOVERNMENT OF WEST BENGAL.**

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT-38/2014

Dated- 26.05.2014

Date of issue: 21.06.2014

The following amendment has been made in the tender document,

Amendment III (Revision of Technical Specification)

SCHEDULE - III

Revised Technical Specification of Electrolyte Analyser

- a) **MEASURED PARAMETERS:** Na⁺, K⁺, Ca⁺⁺, Cl⁻
- b) Should display all results in print out.
- c) It should have inbuilt printer. Analysis time should not be more than 120 seconds.
- d) System should be based on liquid calibration technology.
- e) Should have both auto and manual calibration.
- f) System should be electrode based. System should not run on single test cartridge pack of reagent.
- g) Should work on whole blood, urine and other body fluids.
- h) Should have syringe and capillary sampling
- i) Sample volume should be less than 150 micro litre

- j) Should have numeric/ touch keypad, LCD display and printer
- k) Analyser with memory of 200 tests
- l) System should be supplied complete with all standard accessories, electrodes and start up kits
- m) Onboard life of reagents should not be less than two months
- n) Should supply Reagent pack size of 100T, 200T, 300T, 400T and 500T etc. (as applicable or equivalent amount in ml) based on the demand of the end user. **Bidder should quote cost for each pack size of reagents.**
- o) Accessories: Should be supplied Free of Cost with each equipment
 - i) Heparinised capillary tube – 1000 nos.
 - ii) Printer paper Quantity – Capable to print 4000 test report (equivalent number of rolls, etc.)
- p) Should have safety certificate from a competent authority CE and US FDA
- q) Warranty should be covered for 5 years
- r) CMC cost to be quoted for 5 years
- s) Warranty and CMC should include electrodes, consumables including membrane, lamp, tubing and any other spares / accessories / consumables required for operation of the equipment **only except reagents.**
- t) Bidder should quote the rate of cost per reportable test result presuming 10 (ten) number of tests per day, heparinised capillary tube, complete set of electrodes – these rates will be taken into financial bid evaluation.

u) Labelling & Packing:

The equipment should have a sticker on it with the following information:

- i) Procured by: WBMSCL Ltd.
- ii) Serial Number:
- iii) PMS Done:
- iv) PMS Due:
- v) Toll Free No.:
- vi) Mobile No.:
- vii) Email:
- viii) Facility Asset No.:
- ix) Warranty upto:
- x) CMC Starts on:
- xi) CMC Valid upto:
- xii) Approved CMC Rate per annum:
- xiii) Service Engineer Contact detail: