



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340307/320
E mail: procurement@wbmsc.gov.in

Supply and Commissioning of 5(Five) Different medical equipment at different Hybrid CCUs
across the state of West Bengal
(Submission of Bid through *online*)

2nd call of Bid Reference No.: WBMSCL/NIT-539/2023; dated-03.10.2023

Bid Reference No.: WBMSCL/NIT-264/2024

Dated-06.05.2024

Schedule-III

Amendment-I

Nebulizer

Specification	Specification Name	Bid Requirement (Allowed Values)
GENERAL	Weight (kg)	Less than 1 Kg
FUNCTIONAL	Air Flow in Liter Per Minute (LPM)	5
	Noise level (db)	0.0 - 55.0 Or lower
	Size of uniform Particle (MMAD), (micron)	0.5 - 5.0 Or lower
	Technology	Aerosol with piston electro compressor type
	Nabulizing Rate	0.2 to 0.5 ml/minute for 10 LMP
CONSTRUCTION	Air tube and Air Filter and kit with mask (child and adult), (minimum 5 nos of air filters along with each equipment)	Yes
ELECTRICAL	Rechargeable battery 12 Volt/150AH Lead Acid SMF Type	No
	Moulded Plug power cord length, ISI Marked (Min)	2.0
	1 Hour Back Up of Battery	Yes/No

Specification	Specification Name	Bid Requirement (Allowed Values)
Certificate		European CE(4 digit notified body)/CE-IVD/US FDA/BIS approved