

Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited Swasthya Sathi GN-29, Salt Lake, Sector-V

Kolkata-700091

Phone No (033) 40340307/320 E mail: procurement@wbmsc.gov.in

Procurement of Medical Items for the Blood Centers at Hospitals and Medical Colleges of Govt. of West Bengal (Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT-175/2025 Dated-01.03.2025 2nd call of bid reference no. WBMSCL/NIT-193/2024 Dated-11.03.2024

Amendment-I

Revised Technical Specification

<u>ltem-2</u>

Blood Collection Monitor

- a. The system is used to collect donated blood from the donor at the same time mixing the blood for quality collection of blood.
- b. It is meant for stationary and mobile use. Gentle end to end mixing and control of collection time to give high quality blood suitable for all blood bags.
- c. Volume Setting: Pre-selection of volume to be collected. Tarring of bag volume before collection. Tarring range: 0 600 g. Automatic storage and recall of set volume. Measure volume with best accuracy.
- d. LED indication on commencement of collection.
- e. LED indication and audible alarm at the end of collection.
- f. Indication of time taken for collection.
- g. Indication of blood flow with audio alarm when blood flow is higher or lower than desired.
- h. Continuous display of collected volume, flow and time during collection.
- i. Automatic clamping at termination of preset volume collection.
- j. Automatic release of bag when lifted.
- k. Continuous agitation of blood bags during collection: 12 16 rpm (minimum).

- I. Items carry case for Blood Collection Monitor should be provided for portability.
- m. Should operate on mains as well as rechargeable battery. On battery it should operate for a minimum of 5 8 hours.
- n. The unit shall be capable of operating continuously in ambient temperature of 10 $\,$ 40 $^{0}{\rm C}$ and relative humidity of 15 90%
- o. Power input to be 220 240 VAC, 50Hz Single Phase as appropriate fitted with Indian plug.
- p. Resettable over current breaker shall be fitted for protection.
- q. Warranty : 2 years
- r. Certifications:

Valid CDSC0 Certificate License for the manufacturer(s) & Registration Certificate from CDSC0 for the importer(s) as applicable.

Note:

- 1. Installation Quality (IQ), Operational Quality (OQ), Performance Quality (PQ) certificate to be submitted during installation of the machine.
- 2. 3rd party calibration certification should be also be provided during installation