



## Notice Inviting e-Tender

**West Bengal Medical Services Corporation Limited  
Swasthya Sathi  
GN-29, Salt Lake, Sector-V  
Kolkata-700091**

Phone No (033) 40340307/320  
E mail: procurement@wbsmc.gov.in

**Supply and Commissioning of Anterior Vitrectomy Machine for Dept. of Ophthalmology at Bankura Sammilani Medical College and Hospital, Bankura**

(Submission of Bid through *online*)

**Bid Reference No.: WBMSCL/NIT-1051/2025  
20.12.2025**

**Dated-**

### Amendment I

## Revised Technical Specification

### Anterior Vitrectomy Machine

- Minimum 3000 Cuts per Minute
- 05-600 mmHg panel set venturi aspiration
- Reusable collection bottle with Fluid level Sensor
- Diathermy/Cautery & Forced Positive Reflux .5 Sets of diathermy chord and forceps must be supplied with the machine.
- Light weight and Portable
- 20G, 23G & 25G Compatible.10 Nos. of 23 G cutters must be supplied with the machine.
- User Friendly minimum 5 inches Touch TFT LCD
- Easy to use foot switch – cutting with suction, only suction & Reflux Function with continuous irrigation
- Disposable or Resusable High Speed vitrectomy cutter is must.

- Pressure input: (6-7 Bar) or (90-105 PSI) – Compatible Air Compressor Unit must be supplied with the machine.
- Electrical power: AC 100-240 Volts with suitable UPS
- 5 sets of Reusable Tubing sets must be supplied.
- 5 nos. of Reusable cassette must be supplied.
- 5 sets of Trocher with Infusion Terminal must be supplied.
- Machine must have stand portable with wheels and incorporated with IV Stand.
- 3 sets of Titanium BI Manual(Irrigation Aspiration) probe 22G must be supplied with the machine.
- Standard warranty : 2(Two) Years

**Standard & Safety Certification** : Valid CDSCO Certificate License for the manufacturer(s) & Registration Certificate from CDSCO for the importer(s) as applicable.