



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340307/320
E mail: procurement@wbmsc.gov.in

Procurement of Medical Items for the Blood Centers at Hospitals and Medical Colleges of Govt. of West Bengal
(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT-175/2025	Dated-01.03.2025
2 nd call of bid reference no. WBMSCL/NIT-193/2024 Dated-11.03.2024	

Amendment-I

Revised Technical Specification

Item-13

Haemoglobinometer

Feature	Specifications
Methodology	Reflectance Photometry
Test Time	<10 seconds
Measurement Range	0-25.6 g/dL, 0-256 g/L
Sample	Whole blood / Finger prick
Sample Volume	10-50µL
Power Source	Rechargeable battery with charger and adapter
Units of Measure	g/dL, g/L
Memory	500 to 1,000 records
Automatic Shut Off	3 minutes after last use
Display Size	45 to 50 x 40 to 50 mm
Operating Conditions	10-40°C; ≤90% RH; For indoor use only.
Strips/ Cuvette Storage Conditions	2-30°C; ≤90% RH
Strip/Cuvette	200 numbers at free of cost with each unit
Lancet	Safety complied lockable disposable, non-reusable (single use) sterile lancet from same OEM should be supplied 2000 nos. with each unit
Calibration	Calibration Certificate has to be done in every 6(six) months and should be provided free of cost during entire warranty period as and when required by the facility.
Warranty	2 years

Certifications:

Valid CDSCO Certificate License for the manufacturer(s) & Registration Certificate from CDSCO for the importer(s) as applicable.

Note:

1. Installation Quality (IQ), Operational Quality (OQ), Performance Quality (PQ) certificate to be submitted during installation of the machine.
2. 3rd party calibration certification should be also be provided during installation