



Notice Inviting Tender

West Bengal Medical Services Corporation Limited

Swasthya Sathi

GN-29, Salt Lake, Sector-V

Kolkata-700091

Phone No (033)40340307/320

Email: procurement@wbmsc.gov.in

Procurement of Orthopantomogram (O.P.G) Machine for the Department of Dentistry,
SNP Hospital, the Annex-II of IPGMER & SSKM Hospital.
(Submission of Bid through online)

Bid Reference No.: WBMSCL/NIT-070/2025

Dated-21.01.2025

Amendment-II

Revised Technical Specification

Orthopantomogram (O.P.G) Machine

1. Fully AERB Approved digitalized unit providing clean and sharp panoramic images with enhanced details in anterior region of dental roots of teeth.
2. Actual pano resolution of 5 lp/mm with various capture modes for wide adult and child arches, TMJ Lat and PA, Sinus Lat/PA views.
3. Focal spot 0.5 mm [IEC60336], with variable exposure times for fast, normal and HD Pano modes. Gray Scale 12-14 bit/min. Tube Voltage/Current 60-99KV/4-10 Ma.
4. Generator output should be 2.5-3 KW or more.
5. Dimensions 990mm ± 100 [Length] 1000mm ± 200 [Width], 2300mm ± 100 [Height].
6. Two years warranty if machine from manufacturers is highly essential. Desktop with DICOM Printer is also needed for digitalized film print outs
7. 30 minutes battery Backup for whole systems including console and the O.P.G Machine.
8. Certificate: The bidder should submit valid CDSCO Certificate / Registration / License for both the manufacturer(s) and importer(s) as applicables.
9. Thyroid Guard (0.25mm thickness) - 2 Pcs

10. Lead jacket (0.25mm thickness) – 2 nos.

11. During Warranty period QA is the sole responsible by the vendor or bidder.