



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited

Swasthya Sathi

GN-29, Salt Lake, Sector-V

Kolkata-700091

Phone No (033) 40340307/320

E mail: procurement@wbmsc.gov.in

Supply and Commissioning of different types of Protective Instruments and Gears for the purpose of AERB Compliance across all the Health Facilities under Govt. of West Bengal

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT-035/2025

Dated-10.01.2025

Amendment-I

REVISED TECHNICAL SPECIFICATION

Page 48

Item No. – 5

Movable protective Shield

- a) It should be mobile and sturdy.
- b) Movable Lead Shield should have Lead equivalence of **1.5 mm (Pb)**.
- c) Adjustable Movable Lead Shield can be set between 36 inches and 60 inches high.
- d) Dimensions of the Lead Shield should be 50 x 50 (cms)
- e) It should be convenient to move in any direction.
- f) CDSCO