

बिड दस्तावेज़ / Bid Document

बिड विवरण / Bid Details	
बिड बंद होने की तारीख/समय / Bid End Date/Time	19-11-2025 20:00:00
बिड खुलने की तारीख/समय / Bid Opening Date/Time	19-11-2025 20:30:00
बिड पेशकश वैधता (बंद होने की तारीख से) / Bid Offer Validity (From End Date)	180 (Days)
मंत्रालय/राज्य का नाम / Ministry/State Name	West Bengal
विभाग का नाम / Department Name	Health And Family Welfare Department West Bengal
संगठन का नाम / Organisation Name	West Bengal Medical Services Corporation (wbmsc) Kolkata
कार्यालय का नाम / Office Name	West Bengal
कुल मात्रा / Total Quantity	1
वस्तु श्रेणी / Item Category	Gas Chromatography - Triple Quadrupole Mass Spectrometry (GC - MS - MS) (Q3)
बिडर का न्यूनतम औसत वार्षिक टर्नओवर (3 वर्षों का) / Minimum Average Annual Turnover of the bidder (For 3 Years)	230 Lakh (s)
उन्हीं/समान सेवा के लिए अपेक्षित विगत अनुभव के वर्ष / Years of Past Experience Required for same/similar service	3 Year (s)
एमएसएमई के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है / MSE Exemption for Years of Experience and Turnover	No
स्टार्टअप के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है / Startup Exemption for Years of Experience and Turnover	No
विक्रेता से मांगे गए दस्तावेज़ / Document required from seller	Experience Criteria, Bidder Turnover, Certificate (Requested in ATC), OEM Authorization Certificate, Additional Doc 1 (Requested in ATC) *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
क्या आप निविदाकारों द्वारा अपलोड किए गए दस्तावेज़ों को निविदा में भाग लेने वाले सभी निविदाकारों को दिखाना चाहते हैं? संदर्भ मेनू है / Do you want to show documents uploaded by bidders to all bidders participated in bid?	Yes (Documents submitted as part of a clarification or representation during the tender/bid process will also be displayed to other participated bidders after log in)

बिड विवरण/Bid Details	
बिड लगाने की समय सीमा स्वतः नहीं बढ़ाने के लिए आवश्यक बिड की संख्या। / Minimum number of bids required to disable automatic bid extension	1
दिनों की संख्या, जिनके लिए बिड लगाने की समय-सीमा बढ़ाई जाएगी। / Number of days for which Bid would be auto-extended	3
ऑटो एक्सटेंशन अधिकतम कितनी बार किया जाना है। / Number of Auto Extension count	1
बिड से रिवर्स नीलामी सक्रिय किया/Bid to RA enabled	Yes
रिवर्स नीलामी योग्यता नियम/RA Qualification Rule	H1-Highest Priced Bid Elimination
बिड का प्रकार/Type of Bid	Two Packet Bid
तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय /Time allowed for Technical Clarifications during technical evaluation	2 Days
निरीक्षण आवश्यक (सूचीबद्ध निरीक्षण प्राधिकरण /जेम के साथ पूर्व पंजीकृत एजेंसियों द्वारा)/Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)	No
मूल्यांकन पद्धति/Evaluation Method	Total value wise evaluation
मध्यस्थता खंड/Arbitration Clause	No
सुलह खंड/Mediation Clause	No

ईएमडी विवरण/EMD Detail

एडवाइजरी बैंक/Advisory Bank	ICICI
ईएमडी राशि/EMD Amount	460000

ईपीबीजी विवरण /ePBG Detail

एडवाइजरी बैंक/Advisory Bank	ICICI
ईपीबीजी प्रतिशत (%) /ePBG Percentage(%)	5.00
ईपीबीजी की आवश्यक अवधि (माह) /Duration of ePBG required (Months).	38

(a). जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित केटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज़ प्रस्तुत करने हैं। एमएसई केटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।/EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this

Policy.

(b). ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए। / EMD & Performance security should be in favour of Beneficiary, wherever it is applicable.

लाभार्थी /Beneficiary :

MANAGING DIRECTOR
WEST BENGAL, Health and Family Welfare Department West Bengal, West Bengal Medical Services Corporation (WBMSC) Kolkata,
(West Bengal Medical Services Corporation)

बोली विभाजन लागू नहीं किया गया/ Bid splitting not applied.

एमएसई खरीद वरीयता/MSE Purchase Preference

एमएसई खरीद वरीयता/MSE Purchase Preference	No
---	----

एमआईआई खरीद वरीयता/MII Purchase Preference

एमआईआई खरीद वरीयता/MII Purchase Preference	No
--	----

1. The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated above in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3-year-old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
2. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM of the product offered in the bid {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU for number of Financial years as indicated above in the bid document before the bid opening date. Copies of relevant contracts and delivery acceptance certificates like CRAC to be submitted along with bid in support of having supplied some quantity during each of the Financial year. In case of bunch bids, the category of primary product having highest value should meet this criterion.
3. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.
4. Reverse Auction would be conducted amongst all the technically qualified bidders except the Highest quoting bidder. The technically qualified Highest Quoting bidder will not be allowed to participate in RA. However, H-1 will also be allowed to participate in RA in following cases:
 - i. If number of technically qualified bidders are only 2 or 3.
 - ii. If Buyer has chosen to split the bid amongst N sellers, and H1 bid is coming within N.
 - iii. In case Primary product of only one OEM is left in contention for participation in RA on elimination of H-1.
 - iv. If L-1 is non-MSE and H-1 is eligible MSE and H-1 price is coming within price band of 15% of Non-MSE L-1
 - v. If L-1 is non-MII and H-1 is eligible MII and H-1 price is coming within price band of 20% of Non-MII L-1

Pre Bid Detail(s)

मूल्य भिन्नता खंड दस्तावेज़/Pre-Bid Date and Time	प्री-बिड स्थान/Pre-Bid Venue
04-11-2025 12:00:00	Conference Hall 2nd Floor WBMSCL

Gas Chromatography - Triple Quadrupole Mass Spectrometry (GC - MS - MS) (1 pieces)

तकनीकी विशिष्टियाँ /Technical Specifications

* जेम कैटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
GAS CHROMATOGRAPH (GC) CONFIGURATION	GC Detectors inclusive in the scope of supply	Flame Ionization Detector (FID), Thermal Conductivity Detector (TCD or Hot Wire Detector), Electron Capture Detector (ECD), Nitrogen-Phosphorus Detector (NPD) / Flame Thermionic Detector (FTD), Flame Photometric Detector (FPD), Photo Ionization Detector (PID), Pulsed Flame Photometric Detector (PFPD)
	"Number of FID in the System - inclusive in the scope of Supply	0, 1, 2
	"Number of TCD in the System - inclusive in the scope of Supply	0, 1, 2
	"Number of ECD in the System - inclusive in the scope of Supply	0, 1, 2
	"Number of NPD/FTD in the System - inclusive in the scope of Supply	0, 1, 2
	"Number of FPD in the System - inclusive in the scope of Supply	0, 1, 2
	"Number of PID in the System - inclusive in the scope of Supply	0, 1, 2
	"Number of PFPD in the System - inclusive in the scope of Supply	0, 1, 2
	Number of Inlets GC can Simultaneously Support	1, 2, 3
	Number of Detectors (Ports) GC can Simultaneously Support	1, 2, 3, 4
	Number of Capillary Columns provided in GC - inclusive in the scope of Supply	0, 1, 2
	Type of Capillary Column	Polar, Non-Polar, Semi-Polar, NA
	"GC Capillary Column Length	0, 30, 60, 100 (meter)

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
PERFORMANCE	"GC Detector Minimum Detection Level (MDL)	< 1.4 pg C/s for FID, < 1.8 pg C/s for FID, < 3.0 pg C/s for FID < 400 pg/ml for TCD, < = 1 microgram/ml for TCD, < = 10 microgram/ml for TCD, < 4.5 fg /sec for ECD, < 5.5 fg /sec for ECD, < 0.1pg N /sec for NPD/FTD, < 0.03 pg P /sec for NPD/FTD, < 100 fg P /sec for FPD, <2.5 pg S /sec for FPD, <5 pg S /sec for FPD, < 45 fg P /sec for FPD
COLUMN OVEN (GAS CHROMATOGRAPH - GC)	Maximum Operating Temperature (GC Oven)	upto 400 deg. C, upto 450 deg. C Or higher
	Number of Oven Ramps / Plateaus for Temperature programming	6/7 or more, 9/10 or more, 20/21 or more
PNEUMATIC CONTROL	Number of Injection Ports	1, 2
	Maximum Injection Port heating Temperature	upto 400 deg. C, upto 450 deg. C Or higher
	Pressure Range	upto100 psi, upto 140 psi Or higher
	Pressure set point Resolution (±)(psi)	0.001, 0.01, 0.1 Or lower
	Type of Sample Injection Port	Split, Splitless, split / Splitless, Programmable Temperature Vaporizer (PTV)
	Split Ratio	6000:1, 7000:1, 7500:1, NA
AUTO INJECTOR / AUTO SAMPLER	Compatible Liquid Auto Sampler - inclusive in the scope of supply	Yes, No
	Capacity of the Liquid Auto Sampler (Number of Vials)	0, 8, 100, 150 Or higher
	Number of Injection Ports in the Liquid Auto Sampler	0, 1, 2
	Number of Channels in the Liquid Auto Sampler	0, 2
	"Liquid Auto Sampler Features (Select applicable features Only)"	Fully automated Liquid Auto Sampler, Injects with variable speed & varying syringe sizes, Operates through Software (Software inclusive in the scope of supply), NA
HEAD SPACE SAMPLER	Compatible Head Space Sampler- inclusive in the scope of supply	Yes, No
	Capacity of the Head Space Sampler Carousel (number of Vials)	12 or better, NA
	Head Space Sampler Temperature Range	Upto 200 deg. C or better, NA
PYROLYZER SYSTEM	Compatible Pyrolyzer System for pyrolysis of solid and liquid samples- inclusive in the scope of supply	Yes, No

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
MASS SPECTROPHOTOMETER (MS-MS) SYSTEM	Ionization modes	Electron Ionization (EI), EI/PCI/NCI
	Detector Type	Electron Multiplier, Triple-Axis, HED-EM
	Operating modes	Full Scan Mode, Selective Ion Monitoring (SIM) Mode, Both Full Scan & SIM Modes, Selective /Multiple Reaction Monitoring (SRM/MRM)
PERFORMANCE OF THE MASS SPECTROPHOTOMETER (MS-MS) SYSTEM	Sensitivity (S/N Ratio) in full scan	300:1, 1500:1 Or higher
	Sensitivity (S/N Ratio) in SRM/MRM	6000:1, 15000:1, 30000:1 Or higher
GENERIC	Power Supply	230 V $\pm$ 10%, 415 V $\pm$ 10%
	Built-in Display	Yes, No
	Provision to split the sample from single injection to two different detectors and get results for two detectors simultaneously	Yes, No
	Back Flash - inclusive in the scope of supply	Pre Column Back Flush, Post Column Back Flush, Mid Column Back Flush, NA
SALIENT FEATURES OF GC-MS-MS	"Salient Features of GC-MS (Select applicable features Only)"	Automatic leak testing/detection of Pneumatics, Direct setting of split flow rates and ratios, Flow sensor for control & storage of split ratio, Split and Septum Purge gases, Able to direct inject onto packed and wide-bore capillary columns, Oven power turns-off automatically when the lid/door is opened, Over Heat Protection, Display Unit with built-in Key Board, Controlled by External PC System, Programmable ion source temperature, System includes variety of auto tune algorithms to tune the instrument for maximum sensitivity or for specific target compounds, EI source is inert to active compounds and programmable, EI Source have a dual filament design with automatic selection of other filament if one fails during analysis;, NA, Off-axis detection system for better removal of neutrals, Electron Impact ion source with dual filaments temperature up to 350 °C, Emission current should be able to set up to 300 μ A or better, MRM/SRM Transitions should be 800 transitions/sec, Buyer specific columns - DB-fame /equivalent column 60 m x 0.25 mm id x 0.25 micron- 2 nos , DB-1 /equivalent column 60 m x 0.25 mm id x 0.25 micron- 1 nos, Instrument detection limit sensitivity of < 2 fg during installation using 5 fg of OFN with 1 ul injection with Auto sampler

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
LIBRARIES	Libraries for GC-MS	NIST Library for MS - Licensed version - 2017 or latest, MRM data base Library, Library for Pesticides, Library for Environmental, Library for Forensic, Wiley Library, NA
SOFTWARE	"Software Features (Select Applicable Features only)"	Compatible with Windows 7 or higher version operating system, Seamless integration & control of all the GC parameters, Interface with Multiple GCs through the comprehensive real-time monitoring and diagnostic software, Real time notification via advisories and indicators, Counters, Chromatographic attributes, Instrument diagnostics, Access to maintenance information, Access to maintenance, run and event logs, Retention Time - Locking module for analyzing target compounds in complex matrices
ACCESSORIES	Gas Supply inclusive in the scope of supply	Nitrogen, Hydrogen, Argon, Helium, Zero Air, Not offered
	Number of Cylinders of each type supplied - inclusive in the scope of supply	0, 1, 2
	Gas cylinders certified by PESO	Yes, NA for Gas Cylinder not offered
	Regulator - inclusive in the scope of supply	Dual Stage Regulator, NA
	"Other Accessories inclusive in the scope of supply (Select only those Accessories which are part of supply)"	Gas purification panel for all Gases with moisture trap, Hydrocarbon trap & Oxygen trap, Syringe for manual injection x 05 nos, Auto Sampler Vials x 1000 vials with screw cap, Graphite ferrules x 10 nos, Vessel ferrules x 10 nos, Septa for injector x 100 nos, Inlet liner for splits x 4 nos, Inlet liner for split x 4 nos, O-rings x 10 nos. of each type, Column Cutter x 2 nos, Gas tube Cutter x 2 nos, Tool Kit x 1 no., User / Technical / Maintenance Manual, Standard Samples - NIST / IPL Traceable, Direct Injection Port, Splitter, Multishot Pyrolyser, Single-shot Pyrolyser, Syringe for Auto Sampler injection x 10 nos, EI filaments x 5 nos, NA
DESKTOP WORKSTATION	Desktop workstation Suitable to GC-MS - inclusive in the scope of supply	Yes, No
	Graphics Card Memory Size	0, 1, 2, 4 (gigabyte)
ONLINE UPS	UPS Provided in the Scope of Supply	Yes, No
	Rating of online UPS(KVA)	5.0 KVA, 7.5 KVA, 10 KVA, NA
	Backup time of online UPS	0, 60, 90, 120 (minute)
INSTALLATION & COMMISSIONING / WARRANTY / AVAILABILITY OF SPARES / TRAINING	Warranty on equipment	1, 2, 3, 5 (year)
	Number of days onsite training provided	3, 5, 7, 10 (day)

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
CERTIFICATION / TEST REPORT	"Availability of UL/CE Certificate as per EN 61010-1:2010 Safety requirements for electrical equipment for measurement, control, and laboratory use General requirements"	Yes, No
	Conformance to EMC/EMI as per EN 61326-1:2013 Electrical equipment for measurement, control and laboratory use	Yes, No
	Availability of Test Reports from Central Govt / NABL approved / ILAC accredited Lab to prove conformity to the specification	Yes, No

Additional Specification Parameters - Gas Chromatography - Triple Quadrupole Mass Spectrometry (GC - MS - MS) (1 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
comprehensive warranty	3 Years

* Bidders offering must also comply with the additional specification parameters mentioned above.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Abir Kumar Sasmal	700091, Swasthya Sathi, GN-29, Sector V, Salt Lake	1	60

क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें/Buyer Added Bid Specific Terms and Conditions

1. Service & Support

Availability of Service Centres: Bidder/OEM must have a Functional Service Centre in the State of each Consignee's Location in case of carry-in warranty. (Not applicable in case of goods having on-site warranty). If service center is not already there at the time of bidding, successful bidder / OEM shall have to establish one within 30 days of award of contract. Payment shall be released only after submission of documentary evidence of having Functional Service Centre.

2. Generic

Actual delivery (and Installation & Commissioning (if covered in scope of supply)) is to be done at following address

Regional
Food
Testing
Laboratory
Siliguri
.

3. **Forms of EMD and PBG**

Bidders can also submit the EMD with Payment online through RTGS / internet banking in Beneficiary name

West Bengal Medical Services Corporation Ltd.
Account No.
105605003391
IFSC Code
ICIC0001056
Bank Name
ICICI BANK
Branch address
BIDHANNAGAR, SALT LAKE, SECTOR-V
.

Bidder to indicate bid number and name of bidding entity in the transaction details field at the time of on-line transfer. Bidder has to upload scanned copy / proof of the Online Payment Transfer along with bid.

4. **Generic**

Bidders shall quote only those products (Part of Service delivery) in the bid which are not obsolete in the market and has at least 7 years residual market life i.e. the offered product shall not be declared end-of-life by the OEM before this period.

5. **Buyer Added Bid Specific ATC**

Buyer uploaded ATC document [Click here to view the file.](#)

6. **Certificates**

Bidder's offer is liable to be rejected if they don't upload any of the certificates / documents sought in the Bid document, ATC and Corrigendum if any.

7. **Rate Contract**

Dedicated /toll Free Telephone No. for Service Support : BIDDER/OEM must have Dedicated/toll Free Telephone No. for Service Support.

8. **Rate Contract**

Escalation Matrix For Service Support : Bidder/OEM must provide Escalation Matrix of Telephone Numbers for Service Support.

9. **Warranty**

Warranty period of the supplied products shall be 3 years from the date of final acceptance of goods or after completion of installation, commissioning & testing of goods (if included in the scope of supply), at consignee location. OEM Warranty certificates must be submitted by Successful Bidder at the time of delivery of Goods. The seller should guarantee the rectification of goods in case of any break down during the guarantee period. Seller should have well established Installation, Commissioning, Training, Troubleshooting and Maintenance Service group in INDIA for attending the after sales service. Details of Service Centres near consignee destinations are to be uploaded along with the bid.

10. **Generic**

End User Certificate: Wherever Bidders are insisting for End User Certificate from the Buyer, same shall be provided in Buyer's standard format only.

11. **Generic**

Data Sheet of the product(s) offered in the bid, are to be uploaded along with the bid documents. Buyers can match and verify the Data Sheet with the product specifications offered. In case of any unexplained mismatch of technical parameters, the bid is liable for rejection.

12. **Generic**

For hazardous chemical/item, all precautionary measure as per regulation from the point of transportation/ handling/ storage/ safety/ health/ environment to be undertaken/ specified before dispatch. During dispatch, proper symbol for the hazard/ MSDS/ Batch No./ date of manufacturing/ Gross Weight/ Net Weight/ shelf Life etc are to be written/ printed/ pasted on the body of the packing.

13. **OEM**

IMPORTED PRODUCTS: In case of imported products, OEM or Authorized Seller of OEM should have a registered office in India to provide after sales service support in India. The certificate to this effect should be submitted.

14. **Purchase Preference (Centre)**

Indian suppliers of this item are not allowed to participate and/ or compete in procurement by some foreign governments. Bidders / products from such countries are not eligible / not allowed to participate in this bid in terms of clause 1 (d) of Public Procurement (Preference to Make in India) Order, 2017

15. **Generic**

Installation, Commissioning, Testing, Configuration, Training (if any - which ever is applicable as per scope of supply) is to be carried out by OEM / OEM Certified resource or OEM authorised Reseller.

16. **Generic**

Manufacturer Authorization:Wherever Authorised Distributors/service providers are submitting the bid, Authorisation Form /Certificate with OEM/Original Service Provider details such as name, designation, address, e-mail Id and Phone No. required to be furnished along with the bid

17. **Generic**

Non return of Hard Disk: As per Buyer organization's Security Policy,Faulty Hard Disk of Servers/Desktop Computers/ Laptops etc. will not be returned back to the OEM/supplier against warranty replacement.

18. **Generic**

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity up to 25% of the contracted quantity during the currency of the contract at the contracted rates. The delivery period of quantity shall commence from the last date of original delivery order and in cases where option clause is exercised during the extended delivery period the additional time shall commence from the last date of extended delivery period. The additional delivery time shall be $(\text{Increased quantity} \div \text{Original quantity}) \times \text{Original delivery period (in days)}$, subject to minimum of 30 days. If the original delivery period is less than 30 days, the additional time equals the original delivery period. The Purchaser may extend this calculated delivery duration up to the original delivery period while exercising the option clause. Bidders must comply with these terms.

19. **Warranty**

Over and above the normal Warranty terms as per GeM GTC, the successful bidder / OEM shall have to provide Comprehensive Warranty during the entire Standard warranty period as per contract. : The

comprehensive warranty shall be covering the following scope

All spare parts & labour

(Upload an undertaking with the bid confirming compliance by the bidder if Bidder is taking onus of this compliance. In case OEM is taking onus of this compliance, OEM undertaking is to be uploaded along with Bidder undertaking)

20. **Generic**

Products supplied shall be nontoxic and harmless to health. In the case of toxic materials, Material Safety Data Sheet may be furnished along with the material.

21. **Scope of Supply**

Scope of supply (Bid price to include all cost components) : Supply Installation Testing and Commissioning of Goods

22. **Forms of EMD and PBG**

Successful Bidder can submit the Performance Security in the form of Payment online through RTGS / internet banking also (besides PBG which is allowed as per GeM GTC). On-line payment shall be in Beneficiary name

West Bengal Medical Services Corporation Ltd.

Account No.

105605003391

IFSC Code

ICIC0001056

Bank Name

ICICI BANK

Branch address

BIDHANNAGAR, SALT LAKE, SECTOR-V

. Successful Bidder to indicate Contract number and name of Seller entity in the transaction details field at the time of on-line transfer. Bidder has to upload scanned copy / proof of the Online Payment Transfer in place of PBG within 15 days of award of contract.

23. **Warranty**

Successful bidder will have to ensure that adequate number of dedicated technical service personals / engineers are designated / deployed for attending to the Service Request in a time bound manner and for ensuring Timely Servicing / rectification of defects during warranty period, as per Service level agreement indicated in the relevant clause of the bid.

24. **Generic**

Scope of supply includes Training: Number of employees to be trained

5

, Place for Training

at the consignee location

and Duration of training

2

days.

25. **Generic**

Supplier shall ensure that the Invoice is raised in the name of Consignee with GSTIN of Consignee only.

26. **Warranty**

Timely Servicing / rectification of defects during warranty period: After having been notified of the defects / service requirement during warranty period, Seller has to complete the required Service / Rectification within 3 days time limit. If the Seller fails to complete service / rectification with defined time limit, a penalty of 0.5% of Unit Price of the product shall be charged as penalty for each week of delay from the

seller. Seller can deposit the penalty with the Buyer directly else the Buyer shall have a right to recover all such penalty amount from the Performance Security (PBG). Cumulative Penalty cannot exceed more than 10% of the total contract value after which the Buyer shall have the right to get the service / rectification done from alternate sources at the risk and cost of the Seller besides forfeiture of PBG. Seller shall be liable to re-imburse the cost of such service / rectification to the Buyer.

27. Certificates

The bidder is required to upload, along with the bid, all relevant certificates such as BIS licence, type test certificate, approval certificates and other certificates as prescribed in the Product Specification given in the bid document.

28. Generic

The Buyer has an existing set up / inventory of similar products. The offered / supplied product must be compatible with existing system. The bidder has to ensure Compatibility of the supplied items or shall have to include in the supply the necessary hardware / software to make them compatible at no extra cost to the buyer. The details of items with which compatibility is required are as under:

All spare parts & labour

29. Generic

The successful bidder has to supply all essential accessories required for the successful installation and commissioning of the goods supplied. Besides standard accessories as per normal industry practice, following accessories must be part of supply and cost should be included in bid price:

2

.

30. Generic

1. The Seller shall not assign the Contract in whole or part without obtaining the prior written consent of buyer.
2. The Seller shall not sub-contract the Contract in whole or part to any entity without obtaining the prior written consent of buyer.
3. The Seller shall, notwithstanding the consent and assignment/sub-contract, remain jointly and severally liable and responsible to buyer together with the assignee/ sub-contractor, for and in respect of the due performance of the Contract and the Sellers obligations there under.

31. Generic

The seller is required to print logo as per buyer's requirement.

32. Generic

Without prejudice to Buyer's right to price adjustment by way of discount or any other right or remedy available to Buyer, Buyer may terminate the Contract or any part thereof by a written notice to the Seller, if:

- i) The Seller fails to comply with any material term of the Contract.
- ii) The Seller informs Buyer of its inability to deliver the Material(s) or any part thereof within the stipulated Delivery Period or such inability otherwise becomes apparent.
- iii) The Seller fails to deliver the Material(s) or any part thereof within the stipulated Delivery Period and/or to replace/rectify any rejected or defective Material(s) promptly.
- iv) The Seller becomes bankrupt or goes into liquidation.
- v) The Seller makes a general assignment for the benefit of creditors.
- vi) A receiver is appointed for any substantial property owned by the Seller.
- vii) The Seller has misrepresented to Buyer, acting on which misrepresentation Buyer has placed the Purchase Order on the Seller.

33. Generic

While generating invoice in GeM portal, the seller must upload scanned copy of GST invoice and the screenshot of GST portal confirming payment of GST.

अस्वीकरण/Disclaimer

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for [attached categories](#), trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.
15. Buyer added ATC Clauses which are in contravention of clauses defined by buyer in system generated bid template as indicated above in the Bid Details section, EMD Detail, ePBG Detail and MII and MSE Purchase Preference sections of the bid, unless otherwise allowed by GeM GTC.
16. In a category based bid, adding additional items, through buyer added additional scope of work/ additional terms and conditions/or any other document. If buyer needs more items along with the main item, the same must be added through bunching category based items or by bunching custom catalogs or bunching a BoQ with the main category based item, the same must not be done through ATC or Scope of Work.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

All GeM Sellers / Service Providers are mandated to ensure compliance with all the applicable laws / acts / rules including but not limited to all Labour Laws such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972 etc. Any non-compliance will be treated as breach of contract and Buyer may take suitable actions as per GeM Contract.

[यह बिड सामान्य शर्तों के अंतर्गत भी शासित है /This Bid is also governed by the General Terms and Conditions](#)

जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में

भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा।/In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws.

---धन्यवाद/Thank You---