



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340307/320
E mail: procurement@wbmsc.gov.in

Procurement of Different Medical Equipment for Calcutta National Medical College &
Hospital
(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT- 102/2025

Dated-07.02.2025

AMENDMENT-I

Section IV. Schedule of Requirements

Schedule-III

Digital Lense Meter

1. Should measure sphere range from -25 to + 25
2. Should measure cylinder range from 0 to $\pm 10D$
3. Should measure Axis range from 1 to 180 degree in steps of 1 degree.
4. Should measure Prism range from 0 to 10
5. Should measure addition range from 0 to +10D
6. Should have a wave length of 535 to 660nm
7. Should measure lenses like glass lens, hard/soft contact lenses
8. Should have lens size maximum 10mm
 - a. Should have spectacle lens size 20 to 108 mm
9. Should have minimum 3.5 inch LCD/ LED colour monitor with touch screen. Should have memory function too.
10. Should be supplied with thermal printer **including 20 nos. paper roll**/ink type

11. Should operate from 200 to 240v/50Hz Power supply
12. Should have UV tester, PD measuring, Auto progressive Lens measuring, distortion check
13. Should submit valid CDSCO Certificate/ Registration/ License for both the manufacturer(s) and importer(s) as applicable.