



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited

Swasthya Sathi

GN-29, Salt Lake, Sector-V

Kolkata-700091

Phone No (033) 40340307/320

E mail: procurement@wbmsc.gov.in

Procurement of equipment /item for the Sub-Centre(SC), Primary Health Centre(PHC), Block Public Health Unit (BPHU) and Urban Primary Health Unit (UPHC) of the State of West Bengal under 15th Finance Commission

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT-119/2025

Dated-12.02.2025

Amendment-I

REVISED TECHNICAL SPECIFICATION

Page 43

Schedule -I

Foetal Doppler

1. It should be LCD / LED Display (45mm X 22mm or more)
2. It should be ultrasonic emitting frequency: 2 – 2.5 MHz and power: < 5m Watt /cm²
3. Overall sensitivity at the distances 120 – 200 mm from the face of the probe: 90db
4. Automatic FHR test range 50 – 210 BPM or wider (Beat / Minute)
5. Resolution: 1 BPM
6. Audio output : 0.5 – 1 Watt
7. Should be supplied with electric charger & rechargeable battery (Li – ion/alkaline) with 1 set of spare battery.
8. Good quality water proof case for carrying purpose

9. Good quality water proof probe & gel (5 bottles of 200ml or total quantity of 1000ml) should be supplied with the equipment.
10. Warranty includes the entire machine including probe
11. **Standards, Safety & Training:**
 - a) Valid CDSCO Certificate License for the manufacturer(s) & Registration Certificate from CDSCO for the importer(s) as applicable.
 - b) Warranty: 3 years.
 - c) On site comprehensive training for field staff and support services till customer satisfaction with the system.