



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340307/320
E mail: procurement@wbmsc.gov.in

Supply and Commissioning of Mid End Surgical Operating Microscope for Retinal Surgery for Dept. of Ophthalmology at Jalpaiguri Govt. Medical College and Hospital , Jalpaiguri

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT-786/2025
08.09.2025

Dated-

3rd call of Bid Reference Number WBMSCL/NIT-184/2025 Date : 05.03.2025

Amendment I

Revised Technical Specification

Surgical Operating Microscope for Ophthalmology (Anterior & Posterior Segment Surgery)

MICROSCOPE

- Compact microscope body with high quality & complete apochromatic Optics with 1:6 zoom ration. Magnification factor 0.4X to 2.4X;
- Focusing range 50 mm, Objective lens f-200 mm, 65mm diameter.
- Binocular tube : Tiltable tube with integrated image inverter from OEM only. External Inverter or third party inverter not acceptable.
- **Eye Pieces : 10X / 12.5x with -8D to +5D compensator.**
- Deep View : Depth of field management system for optimal depth perception & maximum light transmission.

ILLUMINATION:

- Stereo Coaxial Illumination system for unique detail recognition , high contrast & stability of red reflex even with strongly pigmented decentered and ametropic eye.
- Retina Protection Device and contrast enhancement.
- Integrated 408 nm UV barrier filter and Blue blocking filter;

X Y COUPLING

- X-Y Coupling with minimum 60 mm x 60 mm adjustable range or better;
- Motorized foot-controlled X-Y coupling with automatic re-centering.

SUSPENSION SYSTEM

- 14 function wired/ wireless foot control, Motorized foot-controlled Zoom and focus with re-centering of focusing position through foot control.
- High quality floor stands with long spring balance suspension arm with effective length of 1 metre or more having load bearing capacity of at least 14 Kg or more.
- Stand should have touch screen LCD display with programming facility for setting the speed of XY, Zoom and focus, Foot Pedal.
- Stand should have cold light fiber Optic illumination 12v 100w Halogen lamp /LED within built lamp housing with two lamps, automatic Lamp changeover facility in case of failure of main lamp.

CCTV ATTACHMENT

- 1 CCD HD Camera or better from standard company mandatorily to eliminate any additional cablings.
- Recorder with USB port, 1 TB HDD, HD LCD TV 32"

WIDE ANGLE RETINA VIEWING SYSTEM (from OEM):

- Wide angled non-Contact observation / viewing system (autoclave able).
- Field of viewing 120 deg., Non-contact Lenses 68D & 128D.
- Both Lenses should mounted and Interchangeable during surgery.
- **Standard & Safety Certification** : Valid CDSCO Certificate License for the manufacturer(s) & Registration Certificate from CDSCO for the importer(s) as applicable