



Notice Inviting e-Tender
West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340307/320

E mail: procurement@wbmsc.gov.in

Procurement of different types of Medical Equipment for CTVS and Neuro Surgery Department ,
Anamoy SSWH under Burdwan Medical college and Hospital
 (Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT-292/2026	Dated-13.03.2026
---	-------------------------

2nd call of bid reference no. WBMSCL/NIT-01/2026 Date-02.01.2026

Amendment-I

Revised Technical Specifications

Schedule -I

CRANIOTOMY SET LIST

Instruments	Quantity
Knife handle (for knife no. 11,12,15,24)	Standard length-2, Long length-1, Long length Bayonetshape-1
Needle holder (various length)	Medium2, long 2
Sponge holding forceps	2
Toothed forceps-medium length	2
Plain forceps—medium length	2
Adson toothed forceps	2
Tissue dissecting scissor	Medium length straight 2 ,Medium length curved 2 Short length straight 2
Thread cutting scissor	2
Dura cutting scissor	2
Fishhook with Spring cable-Yasargil type	4
Malleable Copper brain retractor set	8 inch long, various width (4mmto20mm)
Frazier suction cannula set-teardrop-color coded	1
Elevator Watson Cheyne	2
Elevator-MacDonald	2
Periosteum elevator	2
Kerrison punch set (1mm,2mm,3mm)	Each one
Mastoid retractor set	Small 2 nos, medium 2 nos.
Self-retractor set—Beckmann type	2 nos.
Raney clip applicator	2
Raney disposable clips	200
Penfield dissector set	No-1, No-2, No-3.
Curette bone Volkmann oval cups-Double ended	Small 1,medium 1
Fork tumor Yasargil	3 nos of diffents sizes
Arachnoid knife	2
Brain tumor biopsy	3mm. 5mm, 7mm

MICRO SCISSOR SET FOR NEURO SURGERY

INSTRUMENT	DESCRIPTION	QUANTITY
Micro scissor Bayonet shape	6.5 inch long, straight pointed tip	1
Micro scissor Bayonet shape	6.5 inch long, curved pointed tip	1
Micro scissor Bayonet shape	9 inch long, straight pointed tip	1
Micro scissor Bayonet shape	9 inch long, dissecting curved blunt tip	1

INSTRUMENT	DESCRIPTION	QUANTITY
NEEDLEHOLDER-BAYONET	6 inch long, fine tip	1
NEEDLEHOLDER-BAYONET	8 inch long, fine tip	1

NEEDLEHOLDER—MICROBAYONET

FORCEPS-MICROBAYONET

INSTRUMENT	DESCRIPTION	QUANTITY
FORCEPS-MICROBAYONET	7 inch long, plain 0.6mm tip	1
FORCEPS-MICROBAYONET	9 inch long, plain 0.9mm tip	1

RONGEUR-NIBBLERSET

INSTRUMENT	DESCRIPTION	QUANTITY
RONGEUR-NIBBLER —SINGLE ACTION	Adson straight 7mm 8inch long	1
RONGEUR-NIBBLERACTION —SINGLE	Adson curved 7mm 8inch long	1
RONGEUR-NIBBLERACTION —SINGLE	8inch long, Lempert straight fine tip	1
RONGEUR-NIBBLERACTION —DOUBLE	18 CM long, Ruskin curved tip	1
RONGEUR-NIBBLERACTION —DOUBLE	18 CM long, Ruskin angled tip	1

VP SHUNT SURGERY SET

Mosquito Artery forceps-Curved	6
Mosquito Artery forceps-straight	6
Langenbeck handheld retractor set	Large 1 set, medium 1 set, Pediatric size 1 set
Shunt tunneling forceps	12inch long, serrated tip straight
Shunt tunneling forceps	18 inch long, serrated tip straight
Shunt tube passer with blunt obturator	Adult size 1, Pediatric size 1
Brain ventricular cannula	Various diameters-two numbers different sizes

BURRHOLE SET		
SLNO.	INSTRUMENTS	QUANTITY
1	Sponge holder(local-1,stainless-1)	2
2	Towel clip(paul-3,local-1)	4
3	Mosquito forceps curved	4
4	Mosquito forceps	4
5	Arterycurved6"	4
6	Arterystraight6"	4
7	Allies6"	4
8	Needle holder6"	2
9	Mastoid Retractor(<i>small</i>)	3
10	Dissectingscissor7"curved	2
11	St. scissor (heavy 8"-1, fine 8"-1)	2
12	BP handle(no.3-1,,no.4-1)	2
13	Toothforceps6"	2
14	Brain cannula	1
16	Duraltooth6"	1
17	Plainforceps6"	1
18	Bayonet forceps	1
19	Suction tip (no.6-1,no.10-1,no.-12-1)	3
20	Dissectorno.1	1
21	Dissectorno.4	1
22	Howarth elevator	1
23	Periosteum small	1
24	Kidney dish	2
25	Galipot	2

General Requirements for HAND INSTRUMENTS-

1. All instruments must be manufactured from high-grade surgical stainless steel.
2. Instruments should be corrosion-resistant, autoclavable at 134°C, and compatible with standard sterilization protocols.
3. The surface finish must be matte or satin to reduce glare under surgical lighting.
4. All instruments should carry a manufacturer's mark, batch number.
5. All instruments should Compliant with Indian regulatory standards (CDSCO registration where applicable)

Certificate: The bidder/OEM should submit valid CDSCO Certificate/ Registration/ License for both the manufacturer(s) and importer(s) as applicable