

Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited Swasthya Sathi GN-29, Salt Lake, Sector-V Kolkata-700091

Phone No (033) 40340307/320 E mail: procurement@wbmsc.gov.in

Supply and Commissioning of 1 (one) unit of Electric powered Neuro-Spine Drill for the Department of Neurosurgery of NRS MCH

(Submission of Bid through online)

Bid Reference No.: WBMSCL/NIT-1014/2025 Dated-10.12.2025

AMENDMENT-I

Revised Technical Specifications

Of Electric powered Neuro-Spine Drill

- 1. The system should have minimum torque level from 3.5 Ncm or more.
- 2. The system should have electrically powered high speed motor system with power supply voltage range from 220V to 240V with permitted voltage fluctuation of plus or minus 10%.
- 3. The system should have touch screen control system.
- 4. The system should have universal high-speed motor to connect drill attachments, Craniotomy attachments and optional for micro saw attachments.
- 5. The various parameter should be able to adjust either from touch screen panel or from the multifunctional foot control.
- 6. The system should have wired/wireless multifunctional ergonomically designed foot control system where surgeons should able to control the speed, mode and reverse and forward action.
- 7. The high-speed motor and cable should be compatible with thermo washing machines and autoclave (135 degree Celsius or 383-degree Fahrenheit.)
- 8. The system should have quick release and lock system for tools and attachment.
- 9. No lubrication or seal should be required to run the motor.
- 10. Neuro and Spine application should be possible from a single electrical console.

- 11. The system should have water resistant foot control system with IPXO rating.
- 12. The system should have option for cleaning and maintenance of drill system.

Angled attachment:

- 1. The angled attachment short 7 to 10 cm and long 14 to 20 cm 1 each
- 2. Should be supplied with corresponding burr of sized 2 mm. to 5 mm. for both cutting and diamond 5 each

Straight attachment:

- 1. The straight attachment short 7 to 10 cm and long 14 to 20 cm 1 each
- 3. Should be supplied with corresponding burr of sized 2 mm. to 5 mm. for both cutting and diamond—5 each

Craniotomy Attachments:

1. The Craniotomy attachment of sizes Paediatric – 1 pc. and Adult – 2 pcs.

Bur Hole Attachment:

1. The Bur Hole attachment size should be minimum of 14mm and should have ergonomic shape for fast - 2 pcs. and efficient cranial perforation - 5 pcs.

High Speed Motor:

- 1. The High speed motor should have brushless electric motors.
- 2. The High speed motor should have minimum 75000 rpm with integrated 3.5 m cable autoclavable<134 deg C

One Unit Contain: (Should Supply):

- 1. Touch Screen Control Unit -1 No. and it should have 1 or more universal ports for hand piece and foot switch
- 2. High Speed Motor with Cable 1 No.
- 3. Wired/Wireless Foot Control-1 No.
- 4. Angled attachment Medium & Large attachment- 2 Nos.
- 5. Straight attachment Small & Medium attachment- 2 Nos.
- 6. Craniotome Cutter Medium -10 Nos. (Adult and Paediatric)
- 7. Spine Cutter Medium -10 Nos. (Adult)
- 8. Sterilization Tray -1 No. (SS304 and reputed manufacture)

Terms and Condition:

1. All items should have same manufacturer except Sterilization Tray.

The bidder should submit valid CDSCO certificate/Registration/License for the both Manufacturer and Importers as applicable.